

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042222\
 Data File : VY008465.D
 Acq On : 22 Apr 2022 15:24
 Operator : KP/MD
 Sample : N2526-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-3S

Quant Time: Apr 23 01:53:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.789	168	2801	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	5049	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	5081	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	1623	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	1912	58.015	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.020%
35) Dibromofluoromethane	7.716	113	1442	44.253	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.500%
50) Toluene-d8	10.179	98	6152	47.489	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	94.980%
62) 4-Bromofluorobenzene	12.489	95	1843	42.003	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	84.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042222\
 Data File : VY008465.D
 Acq On : 22 Apr 2022 15:24
 Operator : KP/MD
 Sample : N2526-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-3S

Quant Time: Apr 23 01:53:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

