

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042222\
 Data File : VY008454.D
 Acq On : 22 Apr 2022 09:59
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 23 01:49:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	125	0.00
2 T	Dichlorodifluoromethane	0.454	0.434	4.4	120	0.00
3 P	Chloromethane	0.714	0.715	-0.1	119	0.00
4 C	Vinyl Chloride	0.928	0.942	-1.5#	119	0.00
5 T	Bromomethane	0.863	0.811	6.0	121	0.00
6 T	Chloroethane	0.618	0.625	-1.1	121	0.00
7 T	Trichlorofluoromethane	0.962	0.928	3.5	121	0.00
8 T	Diethyl Ether	0.331	0.317	4.2	114	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.573	0.569	0.7	126	0.00
10 T	Methyl Iodide	0.684	0.715	-4.5	116	0.00
11 T	Tert butyl alcohol	0.068	0.056	17.6	114	0.00
12 CM	1,1-Dichloroethene	0.540	0.538	0.4#	121	0.00
13 T	Acrolein	0.063	0.062	1.6	120	0.00
14 T	Allyl chloride	0.705	0.724	-2.7	124	0.00
15 T	Acrylonitrile	0.169	0.158	6.5	110	0.00
16 T	Acetone	0.137	0.133	2.9	110	0.00
17 T	Carbon Disulfide	1.634	1.634	0.0	120	0.00
18 T	Methyl Acetate	0.384	0.343	10.7	111	0.00
19 T	Methyl tert-butyl Ether	1.710	1.662	2.8	114	0.00
20 T	Methylene Chloride	0.864	0.644	25.5#	104	0.00
21 T	trans-1,2-Dichloroethene	0.659	0.650	1.4	119	0.00
22 T	Diisopropyl ether	1.704	1.673	1.8	116	0.00
23 T	Vinyl Acetate	1.168	1.164	0.3	114	0.00
24 P	1,1-Dichloroethane	1.060	1.055	0.5	119	0.00
25 T	2-Butanone	0.212	0.203	4.2	110	0.00
26 T	2,2-Dichloropropane	0.982	1.011	-3.0	124	0.00
27 T	cis-1,2-Dichloroethene	0.754	0.755	-0.1	119	0.00
28 T	Bromochloromethane	0.428	0.391	8.6	103	0.00
29 T	Tetrahydrofuran	0.137	0.129	5.8	111	0.00
30 C	Chloroform	1.165	1.165	0.0	118	0.00
31 T	Cyclohexane	0.967	0.940	2.8	126	0.00
32 T	1,1,1-Trichloroethane	1.020	1.051	-3.0	124	0.00
33 S	1,2-Dichloroethane-d4	0.588	0.551	6.3	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	121	0.00
35 S	Dibromofluoromethane	0.323	0.307	5.0	101	0.00
36 T	1,1-Dichloropropene	0.513	0.522	-1.8	123	0.00
37 T	Ethyl Acetate	0.284	0.269	5.3	115	0.00
38 T	Carbon Tetrachloride	0.533	0.553	-3.8	124	0.00
39 T	Methylcyclohexane	0.657	0.693	-5.5	126	0.00
40 TM	Benzene	1.498	1.529	-2.1	120	0.00
41 T	Methacrylonitrile	0.138	0.137	0.7	134	0.00
42 TM	1,2-Dichloroethane	0.451	0.447	0.9	114	0.00
43 T	Isopropyl Acetate	0.522	0.517	1.0	113	0.00
44 TM	Trichloroethene	0.413	0.424	-2.7	124	0.00
45 C	1,2-Dichloropropane	0.360	0.368	-2.2#	117	0.00
46 T	Dibromomethane	0.252	0.250	0.8	113	0.00
47 T	Bromodichloromethane	0.535	0.546	-2.1	117	0.00
48 T	Methyl methacrylate	0.222	0.220	0.9	112	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042222\
 Data File : VY008454.D
 Acq On : 22 Apr 2022 09:59
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 23 01:49:34 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	119	0.00
50 S	Toluene-d8	1.176	1.116	5.1	100	0.00
51 T	4-Methyl-2-Pentanone	0.275	0.270	1.8	112	0.00
52 CM	Toluene	0.990	1.017	-2.7#	121	0.00
53 T	t-1,3-Dichloropropene	0.574	0.594	-3.5	116	0.00
54 T	cis-1,3-Dichloropropene	0.636	0.650	-2.2	118	0.00
55 T	1,1,2-Trichloroethane	0.344	0.343	0.3	114	0.00
56 T	Ethyl methacrylate	0.444	0.453	-2.0	113	0.00
57 T	1,3-Dichloropropane	0.561	0.557	0.7	112	0.00
58 T	2-Chloroethyl Vinyl ether	0.243	0.240	1.2	109	0.00
59 T	2-Hexanone	0.189	0.192	-1.6	113	0.00
60 T	Dibromochloromethane	0.395	0.407	-3.0	115	0.00
61 T	1,2-Dibromoethane	0.340	0.341	-0.3	115	0.00
62 S	4-Bromofluorobenzene	0.435	0.402	7.6	101	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	115	0.00
64 T	Tetrachloroethene	0.371	0.386	-4.0	126	0.00
65 PM	Chlorobenzene	1.131	1.171	-3.5	119	0.00
66 T	1,1,1,2-Tetrachloroethane	0.417	0.438	-5.0	120	0.00
67 C	Ethyl Benzene	1.962	2.074	-5.7#	123	0.00
68 T	m/p-Xylenes	0.786	0.836	-6.4	123	0.00
69 T	o-Xylene	0.757	0.799	-5.5	122	0.00
70 T	Styrene	1.275	1.343	-5.3	117	0.00
71 P	Bromoform	0.262	0.273	-4.2	116	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	115	0.00
73 T	Isopropylbenzene	3.791	4.119	-8.7	125	0.00
74 T	N-amyl acetate	0.937	1.003	-7.0	114	0.00
75 P	1,1,2,2-Tetrachloroethane	0.847	0.881	-4.0	113	0.00
76 T	1,2,3-Trichloropropane	0.647	0.537	17.0	93	0.00
77 T	Bromobenzene	0.925	0.966	-4.4	117	0.00
78 T	n-propylbenzene	4.547	4.997	-9.9	125	0.00
79 T	2-Chlorotoluene	2.559	2.729	-6.6	121	0.00
80 T	1,3,5-Trimethylbenzene	3.147	3.370	-7.1	124	0.00
81 T	trans-1,4-Dichloro-2-butene	0.279	0.313	-12.2	119	0.00
82 T	4-Chlorotoluene	2.652	2.795	-5.4	118	0.00
83 T	tert-Butylbenzene	2.732	2.961	-8.4	125	0.00
84 T	1,2,4-Trimethylbenzene	3.122	3.380	-8.3	124	0.00
85 T	sec-Butylbenzene	4.122	4.534	-10.0	126	0.00
86 T	p-Isopropyltoluene	3.452	3.799	-10.1	127	0.00
87 T	1,3-Dichlorobenzene	1.851	1.929	-4.2	120	0.00
88 T	1,4-Dichlorobenzene	1.834	1.902	-3.7	120	0.00
89 T	n-Butylbenzene	3.134	3.486	-11.2	125	0.00
90 T	Hexachloroethane	0.621	0.706	-13.7	126	0.00
91 T	1,2-Dichlorobenzene	1.646	1.678	-1.9	117	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.127	0.132	-3.9	114	0.00
93 T	1,2,4-Trichlorobenzene	0.916	0.955	-4.3	118	0.00
94 T	Hexachlorobutadiene	0.447	0.480	-7.4	125	0.00
95 T	Naphthalene	2.077	2.138	-2.9	113	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042222\
Data File : VY008454.D
Acq On : 22 Apr 2022 09:59
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Apr 23 01:49:34 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 21 14:28:40 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.802	0.832	-3.7	116	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 5