

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042222\
 Data File : VY008468.D
 Acq On : 22 Apr 2022 16:34
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 04/25/2022
 Supervised By :Mahesh Dadoda 04/25/2022

Quant Time: Apr 23 01:53:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	99086	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	170183	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	164704	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	86016	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	58899	50.519	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 101.040%			
35) Dibromofluoromethane	7.716	113	57602	52.445	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 104.900%			
50) Toluene-d8	10.179	98	211347	48.360	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 96.720%			
62) 4-Bromofluorobenzene	12.483	95	76501	51.727	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 103.460%			
Target Compounds					Qvalue	
2) Dichlorodifluoromethane	1.906	85	42177	46.874	ug/l	99
3) Chloromethane	2.113	50	74857	52.869	ug/l	94
4) Vinyl Chloride	2.253	62	100959	54.875	ug/l	98
5) Bromomethane	2.650	94	85604	50.041	ug/l	96
6) Chloroethane	2.790	64	66531	54.357	ug/l	99
7) Trichlorofluoromethane	3.125	101	94973	49.832	ug/l	92
8) Diethyl Ether	3.527	74	31105	47.419	ug/l	71
9) 1,1,2-Trichlorotrifluo...	3.899	101	57449	50.636	ug/l	94
10) Methyl Iodide	4.088	142	67384	45.245	ug/l	90
11) Tert butyl alcohol	4.954	59	23842	196.720	ug/l	97
12) 1,1-Dichloroethene	3.875	96	52925	49.412	ug/l	83
13) Acrolein	3.729	56	21924	176.626	ug/l	99
14) Allyl chloride	4.478	41	70561	50.471	ug/l #	88
15) Acrylonitrile	5.155	53	72950	217.507	ug/l	98
16) Acetone	3.942	43	53410	196.683	ug/l #	85
17) Carbon Disulfide	4.192	76	159237	49.174	ug/l	100
18) Methyl Acetate	4.478	43	31500	41.363	ug/l #	82
19) Methyl tert-butyl Ether	5.216	73	158469	46.758	ug/l	95
20) Methylene Chloride	4.716	84	69574	49.160	ug/l #	76
21) trans-1,2-Dichloroethene	5.216	96	65416	50.067	ug/l #	81
22) Diisopropyl ether	6.118	45	168389	49.875	ug/l #	93
23) Vinyl Acetate	6.057	43	561526	242.592	ug/l #	87
24) 1,1-Dichloroethane	6.015	63	106351	50.605	ug/l	98
25) 2-Butanone	6.984	43	90429	215.250	ug/l #	78
26) 2,2-Dichloropropane	6.978	77	101011	51.895	ug/l	94
27) cis-1,2-Dichloroethene	6.984	96	76054	50.888	ug/l	84
28) Bromochloromethane	7.332	49	41608	49.082	ug/l #	75
29) Tetrahydrofuran	7.344	42	60219	222.550	ug/l #	75
30) Chloroform	7.502	83	118098	51.151	ug/l	100
31) Cyclohexane	7.783	56	94802	49.463	ug/l	86
32) 1,1,1-Trichloroethane	7.703	97	106294	52.577	ug/l	97
36) 1,1-Dichloropropene	7.917	75	90892	52.025	ug/l	97
37) Ethyl Acetate	7.069	43	43009	44.513	ug/l #	93
38) Carbon Tetrachloride	7.899	117	95133	52.409	ug/l	99
39) Methylcyclohexane	9.185	83	120566	53.897	ug/l #	85
40) Benzene	8.161	78	262898	51.579	ug/l	96

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042222\
 Data File : VY008468.D
 Acq On : 22 Apr 2022 16:34
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Apr 23 01:53:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 04/25/2022
 Supervised By :Mahesh Dadoda 04/25/2022

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.319	41	22451m	47.915	ug/l	
42) 1,2-Dichloroethane	8.234	62	75674	49.313	ug/l	90
43) Isopropyl Acetate	8.270	43	82384	46.361	ug/l #	86
44) Trichloroethene	8.941	130	74393	52.978	ug/l	96
45) 1,2-Dichloropropane	9.215	63	64417	52.629	ug/l	96
46) Dibromomethane	9.307	93	41449	48.336	ug/l	94
47) Bromodichloromethane	9.496	83	94558	51.945	ug/l	97
48) Methyl methacrylate	9.289	41	36792	48.695	ug/l #	76
49) 1,4-Dioxane	9.295	88	10038	928.295	ug/l #	81
51) 4-Methyl-2-Pentanone	10.069	43	219277	234.572	ug/l #	85
52) Toluene	10.246	92	178502	52.956	ug/l	95
53) t-1,3-Dichloropropene	10.465	75	98090	50.247	ug/l	100
54) cis-1,3-Dichloropropene	9.929	75	110128	50.908	ug/l #	84
55) 1,1,2-Trichloroethane	10.642	97	57921	49.488	ug/l	96
56) Ethyl methacrylate	10.508	69	75131	49.716	ug/l #	77
57) 1,3-Dichloropropane	10.788	76	95113	49.782	ug/l	97
58) 2-Chloroethyl Vinyl ether	9.782	63	196991	237.848	ug/l	91
59) 2-Hexanone	10.831	43	150719	234.296	ug/l	82
60) Dibromochloromethane	10.983	129	69719	51.848	ug/l	98
61) 1,2-Dibromoethane	11.087	107	56988	49.230	ug/l	100
64) Tetrachloroethene	10.721	164	64841	53.087	ug/l	99
65) Chlorobenzene	11.514	112	193383	51.927	ug/l	100
66) 1,1,1,2-Tetrachloroethane	11.587	131	71714	52.158	ug/l	98
67) Ethyl Benzene	11.593	91	342861	53.047	ug/l	96
68) m/p-Xylenes	11.703	106	276280	106.716	ug/l	91
69) o-Xylene	12.026	106	132408	53.124	ug/l	91
70) Styrene	12.044	104	225283	53.636	ug/l	94
71) Bromoform	12.209	173	41790	48.499	ug/l #	99
73) Isopropylbenzene	12.331	105	344398	52.808	ug/l	100
74) N-amyl acetate	12.142	43	77844	48.308	ug/l #	82
75) 1,1,2,2-Tetrachloroethane	12.581	83	68234	46.827	ug/l	99
76) 1,2,3-Trichloropropane	12.629	75	47032m	42.250	ug/l	
77) Bromobenzene	12.605	156	78630	49.436	ug/l	94
78) n-propylbenzene	12.672	91	416139	53.197	ug/l	98
79) 2-Chlorotoluene	12.757	91	226071	51.348	ug/l	96
80) 1,3,5-Trimethylbenzene	12.812	105	288900	53.369	ug/l	98
81) trans-1,4-Dichloro-2-b...	12.379	75	23099	48.138	ug/l #	83
82) 4-Chlorotoluene	12.855	91	244440	53.577	ug/l	97
83) tert-Butylbenzene	13.074	119	253950	54.040	ug/l	93
84) 1,2,4-Trimethylbenzene	13.117	105	288806	53.770	ug/l	97
85) sec-Butylbenzene	13.251	105	387934	54.701	ug/l	100
86) p-Isopropyltoluene	13.367	119	323403	54.451	ug/l	97
87) 1,3-Dichlorobenzene	13.367	146	163646	51.391	ug/l	100
88) 1,4-Dichlorobenzene	13.446	146	163498	51.832	ug/l	98
89) n-Butylbenzene	13.696	91	300128	55.675	ug/l	100
90) Hexachloroethane	13.958	117	58424	54.714	ug/l	81
91) 1,2-Dichlorobenzene	13.739	146	144503	51.036	ug/l	100
92) 1,2-Dibromo-3-Chloropr...	14.355	75	10043	45.862	ug/l	83
93) 1,2,4-Trichlorobenzene	15.007	180	82078	52.078	ug/l	99
94) Hexachlorobutadiene	15.111	225	40719	52.951	ug/l	98
95) Naphthalene	15.233	128	173162	48.465	ug/l	98
96) 1,2,3-Trichlorobenzene	15.422	180	69396	50.306	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042222\
Data File : VY008468.D
Acq On : 22 Apr 2022 16:34
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022

Quant Time: Apr 23 01:53:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 21 14:28:40 2022
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042222\
 Data File : VY008468.D
 Acq On : 22 Apr 2022 16:34
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/soil
 ALS Vial : 17 Sample Multiplier: 1


Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC

Quant Time: Apr 23 01:53:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Manual Integrations APPROVED


Reviewed By :Krupa Patel 04/25/2022
 Supervised By :Mahesh Dadoda 04/25/2022





#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:168 Resp: 9908
Ion Ratio Lower Upper
168 100
99 53.4 46.9 70.3

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022




#2
Dichlorodifluoromethane
Concen: 46.874 ug/l
RT: 1.906 min Scan# 14
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Tgt Ion: 85 Resp: 42177
Ion Ratio Lower Upper
85 100
87 31.5 16.1 48.2





#3
Chloromethane
Concen: 52.869 ug/l
RT: 2.113 min Scan# 48
Delta R.T. 0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 50 Resp: 7485
Ion Ratio Lower Upper
50 100
52 34.3 24.6 37.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022




#4
Vinyl Chloride
Concen: 54.875 ug/l
RT: 2.253 min Scan# 71
Delta R.T. 0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Tgt Ion: 62 Resp: 100959
Ion Ratio Lower Upper
62 100
64 32.2 25.0 37.6





#5
Bromomethane
Concen: 50.041 ug/l
RT: 2.650 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 94 Resp: 85604
Ion Ratio Lower Upper
94 100
96 96.0 73.5 110.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022



#6
Chloroethane
Concen: 54.357 ug/l
RT: 2.790 min Scan# 159
Delta R.T. -0.006 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34


Tgt Ion: 64 Resp: 66531
Ion Ratio Lower Upper
64 100
66 32.2 25.4 38.0





#7
 Trichlorofluoromethane
 Concen: 49.832 ug/l
 RT: 3.125 min Scan# 214
 Delta R.T. 0.000 min
 Lab File: VY008468.D
 Acq: 22 Apr 2022 16:34




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC




Tgt Ion:101 Resp: 9497
 Ion Ratio Lower Upper
 101 100
 103 65.8 48.0 72.0

Manual Integrations
 APPROVED



Reviewed By :Krupa Patel 04/25/2022
 Supervised By :Mahesh Dadoda 04/25/2022




#8
 Diethyl Ether
 Concen: 47.419 ug/l
 RT: 3.527 min Scan# 280
 Delta R.T. 0.000 min
 Lab File: VY008468.D
 Acq: 22 Apr 2022 16:34



Tgt Ion: 74 Resp: 31105
 Ion Ratio Lower Upper
 74 100
 45 79.6 55.1 165.5









#9
1,1,2-Trichlorotrifluoroethane
Concen: 50.636 ug/l
RT: 3.899 min Scan# 341
Delta R.T. 0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion:101 Resp: 57449
Ion Ratio Lower Upper
101 100
85 45.1 40.6 60.8
151 76.6 64.9 97.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022



#10
Methyl Iodide
Concen: 45.245 ug/l
RT: 4.088 min Scan# 372
Delta R.T. 0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34


Tgt Ion:142 Resp: 67384
Ion Ratio Lower Upper
142 100
127 43.4 41.5 62.3
141 14.1 11.9 17.9





#11
Tert butyl alcohol
Concen: 196.720 ug/l
RT: 4.954 min Scan# 513
Delta R.T. 0.006 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 59 Resp: 2384
Ion Ratio Lower Upper
59 100
57 9.2 6.4 9.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022



#12
1,1-Dichloroethene
Concen: 49.412 ug/l
RT: 3.875 min Scan# 337
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34


Tgt Ion: 96 Resp: 52925
Ion Ratio Lower Upper
96 100
61 148.8 145.8 218.6
98 65.0 52.5 78.7





#13
Acrolein
Concen: 176.626 ug/l
RT: 3.729 min Scan# 313
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 56 Resp: 21924
Ion Ratio Lower Upper
56 100
55 71.2 57.6 86.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022




#14
Allyl chloride
Concen: 50.471 ug/l
RT: 4.478 min Scan# 436
Delta R.T. -0.001 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Tgt Ion: 41 Resp: 70561
Ion Ratio Lower Upper
41 100
39 78.1 58.8 88.2
76 46.8 25.1 37.7#





#15
Acrylonitrile
Concen: 217.507 ug/l
RT: 5.155 min Scan# 548
Delta R.T. -0.006 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 53 Resp: 72950
Ion Ratio Lower Upper
53 100
52 81.7 67.0 100.4
51 37.0 29.8 44.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022



#16
Acetone
Concen: 196.683 ug/l
RT: 3.942 min Scan# 348
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34


Tgt Ion: 43 Resp: 53410
Ion Ratio Lower Upper
43 100
58 35.1 22.0 33.0#





#17
Carbon Disulfide
Concen: 49.174 ug/l
RT: 4.192 min Scan# 389
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 76 Resp: 15923
Ion Ratio Lower Upper
76 100
78 9.1 7.4 11.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022




#18
Methyl Acetate
Concen: 41.363 ug/l
RT: 4.478 min Scan# 436
Delta R.T. 0.006 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Tgt Ion: 43 Resp: 31500
Ion Ratio Lower Upper
43 100
74 29.0 16.6 24.8





#19
Methyl tert-butyl Ether
Concen: 46.758 ug/l
RT: 5.216 min Scan# 51
Delta R.T. -0.006 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 73 Resp: 158469
Ion Ratio Lower Upper
73 100
57 20.8 18.6 27.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022



#20
Methylene Chloride
Concen: 49.160 ug/l
RT: 4.716 min Scan# 475
Delta R.T. 0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34


Tgt Ion: 84 Resp: 69574
Ion Ratio Lower Upper
84 100
49 101.8 114.4 171.6#
51 32.2 35.3 52.9#
86 61.4 53.7 80.5





#21
trans-1,2-Dichloroethene
Concen: 50.067 ug/l
RT: 5.216 min Scan# 511
Delta R.T. 0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 96 Resp: 65410
Ion Ratio Lower Upper
96 100
61 120.0 120.5 180.7
98 67.8 49.8 74.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022




#22
Diisopropyl ether
Concen: 49.875 ug/l
RT: 6.118 min Scan# 705
Delta R.T. 0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Tgt Ion: 45 Resp: 168389
Ion Ratio Lower Upper
45 100
43 60.6 49.3 73.9
87 35.1 19.2 28.8#
59 12.7 8.4 12.6#





#23
 Vinyl Acetate
 Concen: 242.592 ug/l
 RT: 6.057 min Scan# 695
 Delta R.T. -0.001 min
 Lab File: VY008468.D
 Acq: 22 Apr 2022 16:34




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC




Tgt Ion: 43 Resp: 561520
 Ion Ratio Lower Upper
 43 100
 86 13.6 7.1 10.7

Manual Integrations
 APPROVED



Reviewed By :Krupa Patel 04/25/2022
 Supervised By :Mahesh Dadoda 04/25/2022




#24
 1,1-Dichloroethane
 Concen: 50.605 ug/l
 RT: 6.015 min Scan# 688
 Delta R.T. -0.006 min
 Lab File: VY008468.D
 Acq: 22 Apr 2022 16:34



Tgt Ion: 63 Resp: 106351
 Ion Ratio Lower Upper
 63 100
 98 7.9 3.6 10.8
 100 5.0 2.0 5.8





#25

2-Butanone

Concen: 215.250 ug/l

RT: 6.984 min Scan# 847

Delta R.T. 0.000 min

Lab File: VY008468.D

Acq: 22 Apr 2022 16:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 43 Resp: 90429

Ion Ratio Lower Upper

43 100





72 32.5 17.7 26.5

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 04/25/2022

Supervised By :Mahesh Dadoda 04/25/2022



#26

2,2-Dichloropropane

Concen: 51.895 ug/l

RT: 6.978 min Scan# 846

Delta R.T. -0.000 min

Lab File: VY008468.D




Acq: 22 Apr 2022 16:34


Tgt Ion: 77 Resp: 101011

Ion Ratio Lower Upper

77 100


97 23.6 10.4 31.4





#27
 cis-1,2-Dichloroethene
 Concen: 50.888 ug/l
 RT: 6.984 min Scan# 847
 Delta R.T. 0.000 min
 Lab File: VY008468.D
 Acq: 22 Apr 2022 16:34




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC



Tgt Ion: 96 Resp: 76054
 Ion Ratio Lower Upper
 96 100
 61 130.0 0.0 319.8
 98 64.3 0.0 127.0




Manual Integrations
 APPROVED


Reviewed By :Krupa Patel 04/25/2022
 Supervised By :Mahesh Dadoda 04/25/2022



#28
 Bromochloromethane
 Concen: 49.082 ug/l
 RT: 7.332 min Scan# 904
 Delta R.T. -0.006 min
 Lab File: VY008468.D
 Acq: 22 Apr 2022 16:34


Tgt Ion: 49 Resp: 41608
 Ion Ratio Lower Upper
 49 100
 129 2.5 0.0 2.2#
 130 89.6 55.4 83.0#





#29
Tetrahydrofuran
Concen: 222.550 ug/l
RT: 7.344 min Scan# 906
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 42 Resp: 60219
Ion Ratio Lower Upper
42 100
72 54.2 30.3 45.5
71 49.6 29.0 43.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022




#30
Chloroform
Concen: 51.151 ug/l
RT: 7.502 min Scan# 932
Delta R.T. -0.006 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Tgt Ion: 83 Resp: 118098
Ion Ratio Lower Upper
83 100
85 64.5 51.8 77.6





#31
Cyclohexane
Concen: 49.463 ug/l
RT: 7.783 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 56 Resp: 9480
Ion Ratio Lower Upper
56 100
69 36.4 26.7 40.1
84 99.7 67.2 100.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022




#32
1,1,1-Trichloroethane
Concen: 52.577 ug/l
RT: 7.703 min Scan# 965
Delta R.T. -0.001 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34



Tgt Ion: 97 Resp: 106294
Ion Ratio Lower Upper
97 100
99 63.3 50.7 76.1
61 41.7 37.6 56.4





#33

1,2-Dichloroethane-d4

Concen: 50.519 ug/l

RT: 8.142 min Scan# 1037

Delta R.T. 0.000 min

Lab File: VY008468.D

Acq: 22 Apr 2022 16:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDC0050EC

Tgt Ion: 65 Resp: 5889

Ion Ratio Lower Upper

65 100





67 53.7 0.0 99.4

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 04/25/2022

Supervised By :Mahesh Dadoda 04/25/2022



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VY008468.D

Acq: 22 Apr 2022 16:34




Tgt Ion:114 Resp: 170183


Ion Ratio Lower Upper

114 100

63 17.9 0.0 43.2


88 13.6 0.0 31.8





#35
Dibromofluoromethane
Concen: 52.445 ug/l
RT: 7.716 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 113 Resp: 5760
Ion Ratio Lower Upper
113 100
111 102.5 82.6 124.0
192 18.1 17.6 26.4

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022




#36
1,1-Dichloropropene
Concen: 52.025 ug/l
RT: 7.917 min Scan# 1000
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34



Tgt Ion: 75 Resp: 90892
Ion Ratio Lower Upper
75 100
110 37.0 17.1 51.2
77 31.3 24.4 36.6





#37

Ethyl Acetate

Concen: 44.513 ug/l

RT: 7.069 min Scan# 861

Delta R.T. 0.000 min

Lab File: VY008468.D


Acq: 22 Apr 2022 16:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC






Tgt Ion:	43	Resp:	43009
Ion	Ratio	Lower	Upper
43	100		
61	14.7	9.7	14.5
70	11.3	7.1	10.7

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 04/25/2022

Supervised By :Mahesh Dadoda 04/25/2022



#38

Carbon Tetrachloride

Concen: 52.409 ug/l




RT: 7.899 min Scan# 997


Delta R.T. -0.000 min

Lab File: VY008468.D

Acq: 22 Apr 2022 16:34

Tgt Ion:	117	Resp:	95133
Ion	Ratio	Lower	Upper
117	100		
119	97.4	77.9	116.9
121	32.2	24.8	37.2





#39

Methylcyclohexane

Concen: 53.897 ug/l

RT: 9.185 min Scan# 1208

Delta R.T. -0.000 min

Lab File: VY008468.D

Acq: 22 Apr 2022 16:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 83 Resp: 120560

Ion Ratio Lower Upper

83 100

55 62.3 66.5 99.7





98 48.5 37.9 56.9

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 04/25/2022

Supervised By :Mahesh Dadoda 04/25/2022



#40

Benzene

Concen: 51.579 ug/l

RT: 8.161 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VY008468.D




Acq: 22 Apr 2022 16:34


Tgt Ion: 78 Resp: 262898

Ion Ratio Lower Upper

78 100


77 23.5 20.4 30.6





#41
Methacrylonitrile
Concen: 47.915 ug/l m
RT: 7.319 min Scan# 900
Delta R.T. 0.012 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 41 Resp: 2245
Ion Ratio Lower Upper
41 100
39 0.0 92.9 139.3
67 0.0 0.0 0.0
52 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022



#42
1,2-Dichloroethane
Concen: 49.313 ug/l
RT: 8.234 min Scan# 1052
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34


Tgt Ion: 62 Resp: 75674
Ion Ratio Lower Upper
62 100
98 10.9 0.0 14.8





#43
Isopropyl Acetate
Concen: 46.361 ug/l
RT: 8.270 min Scan# 1058
Delta R.T. -0.001 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 82384
Ion Ratio Lower Upper
43 100
61 32.0 20.2 30.2
87 16.3 8.9 13.3




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022



#44
Trichloroethene
Concen: 52.978 ug/l
RT: 8.941 min Scan# 1168
Delta R.T. 0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34

Tgt Ion:130 Resp: 74393
Ion Ratio Lower Upper
130 100
95 90.2 0.0 188.2





#45

1,2-Dichloropropane

Concen: 52.629 ug/l

RT: 9.215 min Scan# 1213

Delta R.T. 0.000 min

Lab File: VY008468.D

Acq: 22 Apr 2022 16:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 63 Resp: 6441

Ion Ratio Lower Upper

63 100





65 30.8 23.0 34.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 04/25/2022

Supervised By :Mahesh Dadoda 04/25/2022



#46

Dibromomethane

Concen: 48.336 ug/l

RT: 9.307 min Scan# 1228

Delta R.T. -0.000 min

Lab File: VY008468.D

Acq: 22 Apr 2022 16:34




Tgt Ion: 93 Resp: 41449


Ion Ratio Lower Upper

93 100

95 82.7 64.3 96.5

174 95.9 83.4 125.0









#47
Bromodichloromethane
Concen: 51.945 ug/l
RT: 9.496 min Scan# 1259
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion: 83 Resp: 94558
Ion Ratio Lower Upper
83 100
85 62.6 51.9 77.9
127 8.3 6.6 10.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022



#48
Methyl methacrylate
Concen: 48.695 ug/l
RT: 9.289 min Scan# 1225
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34

Tgt Ion: 41 Resp: 36792
Ion Ratio Lower Upper
41 100
69 105.5 58.2 87.2#
39 63.2 47.4 71.0





#49

1,4-Dioxane

Concen: 928.295 ug/l

RT: 9.295 min Scan# 1226

Delta R.T. -0.000 min

Lab File: VY008468.D

Acq: 22 Apr 2022 16:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 88 Resp: 1003





Ion	Ratio	Lower	Upper
88	100		
43	28.6	32.8	49.2
58	58.5	59.8	89.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 04/25/2022

Supervised By :Mahesh Dadoda 04/25/2022



#50

Toluene-d8

Concen: 48.360 ug/l

RT: 10.179 min Scan# 1371




Delta R.T. -0.000 min


Lab File: VY008468.D

Acq: 22 Apr 2022 16:34

Tgt Ion: 98 Resp: 211347


Ion	Ratio	Lower	Upper
98	100		
100	65.3	50.2	75.2





#51
4-Methyl-2-Pentanone
Concen: 234.572 ug/l
RT: 10.069 min Scan# 1353
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 43 Resp: 21927
Ion Ratio Lower Upper
43 100
58 44.0 28.1 42.1

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022




#52
Toluene
Concen: 52.956 ug/l
RT: 10.246 min Scan# 1382
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Tgt Ion: 92 Resp: 178502
Ion Ratio Lower Upper
92 100
91 171.3 142.6 213.8





#53
t-1,3-Dichloropropene
Concen: 50.247 ug/l
RT: 10.465 min Scan# 1418
Delta R.T. 0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 75 Resp: 98090
Ion Ratio Lower Upper
75 100
77 32.0 25.4 38.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022




#54
cis-1,3-Dichloropropene
Concen: 50.908 ug/l
RT: 9.929 min Scan# 1330
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Tgt Ion: 75 Resp: 110128
Ion Ratio Lower Upper
75 100
77 32.1 25.6 38.4
39 40.8 47.6 71.4#





#55
1,1,2-Trichloroethane
Concen: 49.488 ug/l
RT: 10.642 min Scan# 1447
Delta R.T. -0.006 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:	97	Resp:	5792
Ion Ratio	Lower	Upper	
97	100		
83	82.4	69.8	104.8
85	52.7	44.3	66.5
99	63.9	49.8	74.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022




#56
Ethyl methacrylate
Concen: 49.716 ug/l
RT: 10.508 min Scan# 1425
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34



Tgt Ion:	69	Resp:	75131
Ion Ratio	Lower	Upper	
69	100		
41	57.0	63.4	95.0#
39	33.1	35.8	53.8#





#57

1,3-Dichloropropane

Concen: 49.782 ug/l

RT: 10.788 min Scan# 1471

Delta R.T. 0.000 min

Lab File: VY008468.D

Acq: 22 Apr 2022 16:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion: 76 Resp: 9511

Ion Ratio Lower Upper

76 100





78 32.7 25.0 37.6

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 04/25/2022

Supervised By :Mahesh Dadoda 04/25/2022



#58

2-Chloroethyl Vinyl ether

Concen: 237.848 ug/l

RT: 9.782 min Scan# 1306

Delta R.T. 0.000 min

Lab File: VY008468.D




Acq: 22 Apr 2022 16:34


Tgt Ion: 63 Resp: 196991

Ion Ratio Lower Upper

63 100


106 31.5 21.5 32.3





#59
2-Hexanone
Concen: 234.296 ug/l
RT: 10.831 min Scan# 1478
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 43 Resp: 150719
Ion Ratio Lower Upper
43 100
58 61.1 24.4 73.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022



#60
Dibromochloromethane
Concen: 51.848 ug/l
RT: 10.983 min Scan# 1503
Delta R.T. 0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34


Tgt Ion:129 Resp: 69719
Ion Ratio Lower Upper
129 100
127 78.8 38.7 116.1





#61
1,2-Dibromoethane
Concen: 49.230 ug/l
RT: 11.087 min Scan# 1520
Delta R.T. 0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:107 Resp: 5698
Ion Ratio Lower Upper
107 100
109 94.7 75.4 113.2

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022




#62
4-Bromofluorobenzene
Concen: 51.727 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Tgt Ion: 95 Resp: 76501
Ion Ratio Lower Upper
95 100
174 83.6 0.0 177.0
176 81.1 0.0 175.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:117 Resp: 164704
Ion Ratio Lower Upper
117 100
82 51.8 43.0 64.4
119 32.0 25.4 38.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022




#64
Tetrachloroethene
Concen: 53.087 ug/l
RT: 10.721 min Scan# 1460
Delta R.T. 0.006 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Tgt Ion:164 Resp: 64841
Ion Ratio Lower Upper
164 100
166 125.9 100.4 150.6
129 91.0 71.4 107.0
131 86.4 70.8 106.2





#65
Chlorobenzene
Concen: 51.927 ug/l
RT: 11.514 min Scan# 11590
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:112 Resp: 19338
Ion Ratio Lower Upper
112 100
114 31.9 25.8 38.6

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022




#66
1,1,1,2-Tetrachloroethane
Concen: 52.158 ug/l
RT: 11.587 min Scan# 1602
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Tgt Ion:131 Resp: 71714
Ion Ratio Lower Upper
131 100
133 95.8 47.9 143.6
119 66.2 30.8 92.3





#67
Ethyl Benzene
Concen: 53.047 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion: 91 Resp: 34286
Ion Ratio Lower Upper
91 100
106 32.6 24.5 36.7

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022




#68
m/p-Xylenes
Concen: 106.716 ug/l
RT: 11.703 min Scan# 1621
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34



Tgt Ion:106 Resp: 276280
Ion Ratio Lower Upper
106 100
91 194.9 166.6 249.8









#69
o-Xylene
Concen: 53.124 ug/l
RT: 12.026 min Scan# 1674
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion:106 Resp: 132408
Ion Ratio Lower Upper
106 100
91 205.9 110.3 331.0




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022



#70
Styrene
Concen: 53.636 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. 0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34


Tgt Ion:104 Resp: 225283
Ion Ratio Lower Upper
104 100
78 45.3 42.4 63.6
103 54.1 44.4 66.6





#71
Bromoform
Concen: 48.499 ug/l
RT: 12.209 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:173 Resp: 41790
Ion Ratio Lower Upper
173 100
175 49.9 24.6 74.0
254 0.0 0.0 0.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.422 min Scan# 1903
Delta R.T. -0.006 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34



Tgt Ion:152 Resp: 86016
Ion Ratio Lower Upper
152 100
115 57.1 28.2 84.7
150 177.3 0.0 347.6





#73

Isopropylbenzene

Concen: 52.808 ug/l

RT: 12.331 min Scan# 11

Delta R.T. -0.000 min

Lab File: VY008468.D

Acq: 22 Apr 2022 16:34

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050EC

Tgt Ion:105 Resp: 344398

Ion Ratio Lower Upper

105 100





120 27.0 13.4 40.2

Manual Integrations

APPROVED

Reviewed By :Krupa Patel 04/25/2022

Supervised By :Mahesh Dadoda 04/25/2022



#74

N-ethyl acetate

Concen: 48.308 ug/l

RT: 12.142 min Scan# 1693

Delta R.T. -0.000 min

Lab File: VY008468.D

Acq: 22 Apr 2022 16:34

Tgt Ion: 43 Resp: 77844




Ion Ratio Lower Upper


43 100

70 51.3 28.7 43.1#

55 25.7 17.3 25.9


61 27.7 17.0 25.6#





#75
1,1,2,2-Tetrachloroethane
Concen: 46.827 ug/l
RT: 12.581 min Scan# 1765
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 83 Resp: 68234
Ion Ratio Lower Upper
83 100
131 10.4 5.6 16.8
85 64.9 32.3 96.8




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022



#76
1,2,3-Trichloropropane
Concen: 42.250 ug/l m
RT: 12.629 min Scan# 1773
Delta R.T. 0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34


Tgt Ion: 75 Resp: 47032
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77
Bromobenzene
Concen: 49.436 ug/l
RT: 12.605 min Scan# 1769
Delta R.T. -0.006 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:156 Resp: 78630
Ion Ratio Lower Upper
156 100
77 178.4 95.5 286.4
158 96.4 48.0 144.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022




#78
n-propylbenzene
Concen: 53.197 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. 0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Tgt Ion: 91 Resp: 416139
Ion Ratio Lower Upper
91 100
120 23.5 11.3 33.9





#79
2-Chlorotoluene
Concen: 51.348 ug/l
RT: 12.757 min Scan# 1794
Delta R.T. -0.001 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 91 Resp: 22607
Ion Ratio Lower Upper
91 100
126 36.2 16.9 50.6




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022



#80
1,3,5-Trimethylbenzene
Concen: 53.369 ug/l
RT: 12.812 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34


Tgt Ion:105 Resp: 288900
Ion Ratio Lower Upper
105 100
120 50.2 24.4 73.4





#81
trans-1,4-Dichloro-2-butene
Concen: 48.138 ug/l
RT: 12.379 min Scan# 1732
Delta R.T. -0.001 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 75 Resp: 23099
Ion Ratio Lower Upper
75 100
53 82.0 85.3 127.9
89 46.9 37.0 55.4




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022



#82
4-Chlorotoluene
Concen: 53.577 ug/l
RT: 12.855 min Scan# 1810
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34


Tgt Ion: 91 Resp: 244440
Ion Ratio Lower Upper
91 100
126 36.0 17.0 51.0





#83
tert-Butylbenzene
Concen: 54.040 ug/l
RT: 13.074 min Scan# 1846
Delta R.T. -0.001 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:119 Resp: 253950
Ion Ratio Lower Upper
119 100
91 62.0 34.8 104.4
134 27.1 13.3 39.9

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022




#84
1,2,4-Trimethylbenzene
Concen: 53.770 ug/l
RT: 13.117 min Scan# 1853
Delta R.T. -0.006 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Tgt Ion:105 Resp: 288806
Ion Ratio Lower Upper
105 100
120 46.5 22.3 66.9





#85
 sec-Butylbenzene
 Concen: 54.701 ug/l
 RT: 13.251 min Scan# 1875
 Delta R.T. 0.000 min
 Lab File: VY008468.D
 Acq: 22 Apr 2022 16:34




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050EC




Tgt Ion:105 Resp: 387934
 Ion Ratio Lower Upper
 105 100
 134 20.4 10.2 30.4

Manual Integrations
 APPROVED



Reviewed By :Krupa Patel 04/25/2022
 Supervised By :Mahesh Dadoda 04/25/2022




#86
 p-Isopropyltoluene
 Concen: 54.451 ug/l
 RT: 13.367 min Scan# 1894
 Delta R.T. 0.000 min
 Lab File: VY008468.D
 Acq: 22 Apr 2022 16:34




Tgt Ion:119 Resp: 323403
 Ion Ratio Lower Upper
 119 100
 134 26.9 13.2 39.5
 91 22.1 12.4 37.2





#87
1,3-Dichlorobenzene
Concen: 51.391 ug/l
RT: 13.367 min Scan# 1894
Delta R.T. 0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:146 Resp: 163640
Ion Ratio Lower Upper
146 100
111 39.3 19.5 58.5
148 63.3 31.6 95.0

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022




#88
1,4-Dichlorobenzene
Concen: 51.832 ug/l
RT: 13.446 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Tgt Ion:146 Resp: 163498
Ion Ratio Lower Upper
146 100
111 37.5 19.3 57.8
148 63.5 32.6 97.8





#89
n-Butylbenzene
Concen: 55.675 ug/l
RT: 13.696 min Scan# 1948
Delta R.T. 0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC



Tgt Ion: 91 Resp: 300128
Ion Ratio Lower Upper
91 100
92 52.9 26.3 78.9
134 26.1 13.0 38.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022



#90
Hexachloroethane
Concen: 54.714 ug/l
RT: 13.958 min Scan# 1991
Delta R.T. -0.001 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34

Tgt Ion:117 Resp: 58424
Ion Ratio Lower Upper
117 100
201 68.4 43.0 128.8









#91
1,2-Dichlorobenzene
Concen: 51.036 ug/l
RT: 13.739 min Scan# 1955
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC

Tgt Ion:146 Resp: 14450
Ion Ratio Lower Upper
146 100
111 40.6 20.3 60.9
148 64.7 32.3 96.9




Manual Integrations
APPROVED


Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022



#92
1,2-Dibromo-3-Chloropropane
Concen: 45.862 ug/l
RT: 14.355 min Scan# 2056
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34


Tgt Ion: 75 Resp: 10043
Ion Ratio Lower Upper
75 100
155 89.4 38.0 113.9
157 113.7 47.8 143.3





#93
1,2,4-Trichlorobenzene
Concen: 52.078 ug/l
RT: 15.007 min Scan# 2163
Delta R.T. 0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:180 Resp: 82078
Ion Ratio Lower Upper
180 100
182 93.7 47.3 141.9
145 30.4 14.6 43.8

Manual Integrations
APPROVED



Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022




#94
Hexachlorobutadiene
Concen: 52.951 ug/l
RT: 15.111 min Scan# 2180
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Tgt Ion:225 Resp: 40719
Ion Ratio Lower Upper
225 100
223 63.9 30.8 92.3
227 63.1 31.6 94.7





#95
Naphthalene
Concen: 48.465 ug/l
RT: 15.233 min Scan# 2111
Delta R.T. -0.006 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050EC




Tgt Ion:128 Resp: 173162
Ion Ratio Lower Upper
128 100
127 12.7 11.0 16.4
129 11.0 8.4 12.6

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 04/25/2022
Supervised By :Mahesh Dadoda 04/25/2022



#96
1,2,3-Trichlorobenzene
Concen: 50.306 ug/l
RT: 15.422 min Scan# 2231
Delta R.T. -0.000 min
Lab File: VY008468.D
Acq: 22 Apr 2022 16:34



Tgt Ion:180 Resp: 69396
Ion Ratio Lower Upper
180 100
182 96.1 47.3 141.9
145 33.0 15.3 45.9

