

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042325\
 Data File : VY021970.D
 Acq On : 23 Apr 2025 14:40
 Operator : SY/MD
 Sample : Q1848-03
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ETGI-275

Quant Time: Apr 24 03:46:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	348450	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	649608	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	561006	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	211773	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	162377	50.450	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	100.900%
35) Dibromofluoromethane	7.634	113	212323	49.473	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.940%
50) Toluene-d8	10.109	98	767271	47.422	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	94.840%
62) 4-Bromofluorobenzene	12.407	95	216280	39.709	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	79.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042325\
Data File : VY021970.D
Acq On : 23 Apr 2025 14:40
Operator : SY/MD
Sample : Q1848-03
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ETGI-275

Quant Time: Apr 24 03:46:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

