

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042325\
 Data File : VY021971.D
 Acq On : 23 Apr 2025 15:03
 Operator : SY/MD
 Sample : Q1848-05
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ETGI-342

Quant Time: Apr 24 03:46:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	332161	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	626870	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	542304	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	203402	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	152938	49.848	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	99.700%
35) Dibromofluoromethane	7.634	113	199330	48.131	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.260%
50) Toluene-d8	10.109	98	743508	47.620	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.240%
62) 4-Bromofluorobenzene	12.408	95	204774	38.960	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	77.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042325\
Data File : VY021971.D
Acq On : 23 Apr 2025 15:03
Operator : SY/MD
Sample : Q1848-05
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ETGI-342

Quant Time: Apr 24 03:46:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

