

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042325\
 Data File : VY021980.D
 Acq On : 23 Apr 2025 18:34
 Operator : SY/MD
 Sample : Q1858-03
 Misc : 3.86g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-3

Quant Time: Apr 24 03:47:50 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	306739	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	573313	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	503025	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	189503	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	149825	52.880	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	105.760%
35) Dibromofluoromethane	7.634	113	192042	50.703	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.400%
50) Toluene-d8	10.103	98	686663	48.088	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.180%
62) 4-Bromofluorobenzene	12.408	95	189650	39.453	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	78.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042325\
Data File : VY021980.D
Acq On : 23 Apr 2025 18:34
Operator : SY/MD
Sample : Q1858-03
Misc : 3.86g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-3

Quant Time: Apr 24 03:47:50 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

