

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY042423\  
 Data File : VY013406.D  
 Acq On : 24 Apr 2023 17:26  
 Operator : KP/MD  
 Sample : 02371-01  
 Misc : 5.03g/5.0mL/MSVOA\_Y/SOIL  
 ALS Vial : 18 Sample Multiplier: 1

Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 SB-26-9.0-9.5

Quant Time: Apr 25 03:15:18 2023  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y042423S.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 25 02:39:58 2023  
 Response via : Initial Calibration


| Compound                      | R.T.           | QIon | Response             | Conc    | Units  | Dev(Min) |
|-------------------------------|----------------|------|----------------------|---------|--------|----------|
| -----                         |                |      |                      |         |        |          |
| Internal Standards            |                |      |                      |         |        |          |
| 1) Pentafluorobenzene         | 7.789          | 168  | 2301                 | 50.000  | ug/l   | 0.00     |
| 34) 1,4-Difluorobenzene       | 8.691          | 114  | 4578                 | 50.000  | ug/l   | 0.00     |
| 63) Chlorobenzene-d5          | 11.495         | 117  | 5084                 | 50.000  | ug/l   | # 0.00   |
| 72) 1,4-Dichlorobenzene-d4    | 13.422         | 152  | 1451                 | 50.000  | ug/l   | 0.00     |
|                               |                |      |                      |         |        |          |
| System Monitoring Compounds   |                |      |                      |         |        |          |
| 33) 1,2-Dichloroethane-d4     | 8.142          | 65   | 3159                 | 125.238 | ug/l   | 0.00     |
| Spiked Amount 50.000          | Range 50 - 163 |      | Recovery = 250.480%# |         |        |          |
| 35) Dibromofluoromethane      | 7.715          | 113  | 2053                 | 71.301  | ug/l   | 0.00     |
| Spiked Amount 50.000          | Range 54 - 147 |      | Recovery = 142.600%  |         |        |          |
| 50) Toluene-d8                | 10.178         | 98   | 5592                 | 54.593  | ug/l   | 0.00     |
| Spiked Amount 50.000          | Range 58 - 134 |      | Recovery = 109.180%  |         |        |          |
| 62) 4-Bromofluorobenzene      | 12.489         | 95   | 1867                 | 54.066  | ug/l   | 0.01     |
| Spiked Amount 50.000          | Range 30 - 143 |      | Recovery = 108.140%  |         |        |          |
|                               |                |      |                      |         |        |          |
| Target Compounds              |                |      |                      |         |        | Qvalue   |
| 3) Chloromethane              | 2.119          | 50   | 535                  | 18.246  | ug/l # | 59       |
| 5) Bromomethane               | 2.680          | 94   | 463                  | 19.900  | ug/l # | 52       |
| 6) Chloroethane               | 2.814          | 64   | 158                  | 7.834   | ug/l # | 44       |
| 8) Diethyl Ether              | 3.308          | 74   | 30                   | 1.992   | ug/l # | 2        |
| 10) Methyl Iodide             | 4.100          | 142  | 849                  | 27.247  | ug/l # | 64       |
| 12) 1,1-Dichloroethene        | 3.594          | 96   | 47                   | 1.965   | ug/l # | 1        |
| 13) Acrolein                  | 3.960          | 56   | 26                   | 3.051   | ug/l # | 16       |
| 14) Allyl chloride            | 4.503          | 41   | 47                   | 1.059   | ug/l # | 22       |
| 16) Acetone                   | 3.960          | 43   | 318                  | 32.013  | ug/l # | 43       |
| 18) Methyl Acetate            | 4.509          | 43   | 143                  | 4.359   | ug/l # | 51       |
| 20) Methylene Chloride        | 4.722          | 84   | 313                  | 2.349   | ug/l # | 60       |
| 23) Vinyl Acetate             | 6.063          | 43   | 74                   | 1.507   | ug/l # | 72       |
| 25) 2-Butanone                | 6.972          | 43   | 68                   | 5.398   | ug/l # | 47       |
| 29) Tetrahydrofuran           | 7.374          | 42   | 42                   | 5.356   | ug/l # | 32       |
| 31) Cyclohexane               | 7.776          | 56   | 60                   | 1.329   | ug/l # | 14       |
| 36) 1,1-Dichloropropene       | 7.795          | 75   | 248                  | 5.190   | ug/l # | 40       |
| 41) Methacrylonitrile         | 7.356          | 41   | 21                   | 1.253   | ug/l # | 15       |
| 48) Methyl methacrylate       | 9.276          | 41   | 41                   | 1.471   | ug/l # | 12       |
| 51) 4-Methyl-2-Pentanone      | 10.081         | 43   | 62                   | 1.833   | ug/l # | 37       |
| 58) 2-Chloroethyl Vinyl ether | 9.794          | 63   | 29                   | 1.999   | ug/l # | 45       |
| 59) 2-Hexanone                | 10.782         | 43   | 55                   | 2.237   | ug/l   | 75       |
| 60) Dibromochloromethane      | 10.709         | 129  | 52                   | 1.536   | ug/l # | 11       |
| 76) 1,2,3-Trichloropropane    | 12.483         | 75   | 1078                 | 55.089  | ug/l # | 100      |
| 81) trans-1,4-Dichloro-2-b... | 12.483         | 75   | 1078                 | 112.687 | ug/l # | 12       |
| 87) 1,3-Dichlorobenzene       | 13.440         | 146  | 86                   | 1.697   | ug/l # | 63       |
| 88) 1,4-Dichlorobenzene       | 13.440         | 146  | 86                   | 1.680   | ug/l # | 62       |
| 95) Naphthalene               | 15.238         | 128  | 159                  | 2.687   | ug/l # | 69       |
| -----                         |                |      |                      |         |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY042423\  
Data File : VY013406.D  
Acq On : 24 Apr 2023 17:26  
Operator : KP/MD  
Sample : 02371-01  
Misc : 5.03g/5.0mL/MSVOA\_Y/SOIL  
ALS Vial : 18 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
SB-26-9.0-9.5

Quant Time: Apr 25 03:15:18 2023  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y042423S.M  
Quant Title : SW846 8260  
QLast Update : Tue Apr 25 02:39:58 2023  
Response via : Initial Calibration









#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 7.789 min Scan# 91  
Delta R.T. -0.000 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26




Instrument :  
MSVOA\_Y  
ClientSampleId :  
SB-26-9.0-9.5


Tgt Ion:168 Resp: 2301  
Ion Ratio Lower Upper  
168 100  
99 47.8 41.8 62.6



#3  
Chloromethane  
Concen: 18.246 ug/l  
RT: 2.119 min Scan# 49  
Delta R.T. 0.006 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26

Tgt Ion: 50 Resp: 535  
Ion Ratio Lower Upper  
50 100  
52 55.3 25.9 38.9#









#5  
Bromomethane  
Concen: 19.900 ug/l  
RT: 2.680 min Scan# 141  
Delta R.T. 0.024 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26



Instrument :  
MSVOA\_Y  
ClientSampleId :  
SB-26-9.0-9.5


Tgt Ion: 94 Resp: 463  
Ion Ratio Lower Upper  
94 100  
96 46.3 74.2 111.2#



#6  
Chloroethane  
Concen: 7.834 ug/l  
RT: 2.814 min Scan# 163  
Delta R.T. 0.024 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26

Tgt Ion: 64 Resp: 158  
Ion Ratio Lower Upper  
64 100  
66 0.0 24.6 37.0#









#8  
Diethyl Ether  
Concen: 1.992 ug/l  
RT: 3.308 min Scan# 244  
Delta R.T. -0.219 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26




Instrument :  
MSVOA\_Y  
ClientSampleId :  
SB-26-9.0-9.5


Tgt Ion: 74 Resp: 30  
Ion Ratio Lower Upper  
74 100  
45 0.0 48.2 144.6#



#10  
Methyl Iodide  
Concen: 27.247 ug/l  
RT: 4.100 min Scan# 374  
Delta R.T. 0.012 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26

Tgt Ion: 142 Resp: 849  
Ion Ratio Lower Upper  
142 100  
127 21.7 36.1 54.1#  
141 0.0 11.6 17.4#









#12  
1,1-Dichloroethene  
Concen: 1.965 ug/l  
RT: 3.594 min Scan# 291  
Delta R.T. -0.275 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26




Instrument :  
MSVOA\_Y  
ClientSampleId :  
SB-26-9.0-9.5


Tgt Ion: 96 Resp: 47  
Ion Ratio Lower Upper  
96 100  
61 0.0 123.9 185.9#  
98 0.0 48.9 73.3#



#13  
Acrolein  
Concen: 3.051 ug/l  
RT: 3.960 min Scan# 351  
Delta R.T. 0.231 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26

Tgt Ion: 56 Resp: 26  
Ion Ratio Lower Upper  
56 100  
55 0.0 54.1 81.1#









#14  
 Allyl chloride  
 Concen: 1.059 ug/l  
 RT: 4.503 min Scan# 44  
 Delta R.T. 0.025 min  
 Lab File: VY013406.D  
 Acq: 24 Apr 2023 17:26




Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 SB-26-9.0-9.5


Tgt Ion: 41 Resp: 47  
 Ion Ratio Lower Upper  
 41 100  
 39 0.0 56.6 85.0#  
 76 0.0 30.5 45.7#



#16  
 Acetone  
 Concen: 32.013 ug/l  
 RT: 3.960 min Scan# 351  
 Delta R.T. 0.012 min  
 Lab File: VY013406.D  
 Acq: 24 Apr 2023 17:26

Tgt Ion: 43 Resp: 318  
 Ion Ratio Lower Upper  
 43 100  
 58 0.0 25.4 38.0#









#18  
Methyl Acetate  
Concen: 4.359 ug/l  
RT: 4.509 min Scan# 441  
Delta R.T. 0.031 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26




Instrument :  
MSVOA\_Y  
ClientSampleId :  
SB-26-9.0-9.5


Tgt Ion: 43 Resp: 143  
Ion Ratio Lower Upper  
43 100  
74 0.0 19.4 29.2#



#20  
Methylene Chloride  
Concen: 2.349 ug/l  
RT: 4.722 min Scan# 476  
Delta R.T. 0.006 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26

Tgt Ion: 84 Resp: 313  
Ion Ratio Lower Upper  
84 100  
49 176.1 88.0 132.0#  
51 34.9 27.4 41.0  
86 45.4 50.4 75.6#









#23  
 Vinyl Acetate  
 Concen: 1.507 ug/l  
 RT: 6.063 min Scan# 64  
 Delta R.T. 0.006 min  
 Lab File: VY013406.D  
 Acq: 24 Apr 2023 17:26




Instrument :  
 MSVOA\_Y  
 ClientSampleId :  
 SB-26-9.0-9.5


Tgt Ion: 43 Resp: 74  
 Ion Ratio Lower Upper  
 43 100  
 86 0.0 8.5 12.7#



#25  
 2-Butanone  
 Concen: 5.398 ug/l  
 RT: 6.972 min Scan# 845  
 Delta R.T. -0.012 min  
 Lab File: VY013406.D  
 Acq: 24 Apr 2023 17:26

Tgt Ion: 43 Resp: 68  
 Ion Ratio Lower Upper  
 43 100  
 72 0.0 21.9 32.9#





#29

Tetrahydrofuran

Concen: 5.356 ug/l

RT: 7.374 min Scan# 911

Delta R.T. 0.030 min

Lab File: VY013406.D

Acq: 24 Apr 2023 17:26

Instrument :

MSVOA\_Y

ClientSampleId :

SB-26-9.0-9.5





Tgt Ion: 42 Resp: 42

Ion Ratio Lower Upper

42 100

72 0.0 35.9 53.9#

71 0.0 34.3 51.5#



#31

Cyclohexane

Concen: 1.329 ug/l

RT: 7.776 min Scan# 977

Delta R.T. -0.007 min

Lab File: VY013406.D

Acq: 24 Apr 2023 17:26




Tgt Ion: 56 Resp: 60


Ion Ratio Lower Upper

56 100

69 0.0 29.2 43.8#

84 0.0 73.0 109.6#









#33  
1,2-Dichloroethane-d4  
Concen: 125.238 ug/l  
RT: 8.142 min Scan# 1037  
Delta R.T. 0.006 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26




Instrument :  
MSVOA\_Y  
ClientSampleId :  
SB-26-9.0-9.5


Tgt Ion: 65 Resp: 3159  
Ion Ratio Lower Upper  
65 100  
67 52.7 0.0 103.4



#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 8.691 min Scan# 1127  
Delta R.T. -0.000 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26

Tgt Ion: 114 Resp: 4578  
Ion Ratio Lower Upper  
114 100  
63 33.7 0.0 36.6  
88 15.5 0.0 28.8









#35  
Dibromofluoromethane  
Concen: 71.301 ug/l  
RT: 7.715 min Scan# 91  
Delta R.T. -0.001 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26




Instrument :  
MSVOA\_Y  
ClientSampleId :  
SB-26-9.0-9.5


Tgt Ion:113 Resp: 2053  
Ion Ratio Lower Upper  
113 100  
111 111.4 82.2 123.2  
192 21.9 16.7 25.1



#36  
1,1-Dichloropropene  
Concen: 5.190 ug/l  
RT: 7.795 min Scan# 980  
Delta R.T. -0.122 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26

Tgt Ion: 75 Resp: 248  
Ion Ratio Lower Upper  
75 100  
110 0.0 18.7 56.1#  
77 0.0 24.6 37.0#









#41  
Methacrylonitrile  
Concen: 1.253 ug/l  
RT: 7.356 min Scan# 900  
Delta R.T. 0.049 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26




Instrument :  
MSVOA\_Y  
ClientSampleId :  
SB-26-9.0-9.5


|          |       |       |       |
|----------|-------|-------|-------|
| Tgt Ion: | 41    | Resp: | 21    |
| Ion      | Ratio | Lower | Upper |
| 41       | 100   |       |       |
| 39       | 123.8 | 44.2  | 66.4# |
| 67       | 0.0   | 62.9  | 94.3# |
| 52       | 0.0   | 26.4  | 39.6# |



#48  
Methyl methacrylate  
Concen: 1.471 ug/l  
RT: 9.276 min Scan# 1223  
Delta R.T. -0.013 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26

|          |       |       |        |
|----------|-------|-------|--------|
| Tgt Ion: | 41    | Resp: | 41     |
| Ion      | Ratio | Lower | Upper  |
| 41       | 100   |       |        |
| 69       | 0.0   | 70.2  | 105.4# |
| 39       | 0.0   | 47.9  | 71.9#  |









#50  
Toluene-d8  
Concen: 54.593 ug/l  
RT: 10.178 min Scan# 11  
Delta R.T. -0.001 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26




Instrument :  
MSVOA\_Y  
ClientSampleId :  
SB-26-9.0-9.5


Tgt Ion: 98 Resp: 5592  
Ion Ratio Lower Upper  
98 100  
100 67.0 50.5 75.7



#51  
4-Methyl-2-Pentanone  
Concen: 1.833 ug/l  
RT: 10.081 min Scan# 1355  
Delta R.T. 0.012 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26

Tgt Ion: 43 Resp: 62  
Ion Ratio Lower Upper  
43 100  
58 0.0 30.6 46.0#









#58  
2-Chloroethyl Vinyl ether  
Concen: 1.999 ug/l  
RT: 9.794 min Scan# 11  
Delta R.T. 0.012 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26




Instrument :  
MSVOA\_Y  
ClientSampleId :  
SB-26-9.0-9.5


Tgt Ion: 63 Resp: 29  
Ion Ratio Lower Upper  
63 100  
106 0.0 23.5 35.3#



#59  
2-Hexanone  
Concen: 2.237 ug/l  
RT: 10.782 min Scan# 1470  
Delta R.T. -0.049 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26

Tgt Ion: 43 Resp: 55  
Ion Ratio Lower Upper  
43 100  
58 34.5 26.3 78.8









#60  
Dibromochloromethane  
Concen: 1.536 ug/l  
RT: 10.709 min Scan# 1458  
Delta R.T. -0.274 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26




Instrument :  
MSVOA\_Y  
ClientSampleId :  
SB-26-9.0-9.5


Tgt Ion:129 Resp: 52  
Ion Ratio Lower Upper  
129 100  
127 0.0 38.7 116.1#



#62  
4-Bromofluorobenzene  
Concen: 54.066 ug/l  
RT: 12.489 min Scan# 1750  
Delta R.T. 0.012 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26

Tgt Ion: 95 Resp: 1867  
Ion Ratio Lower Upper  
95 100  
174 79.8 0.0 182.0  
176 75.7 0.0 175.4









#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 11.495 min Scan# 11  
Delta R.T. 0.006 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26




Instrument :  
MSVOA\_Y  
ClientSampleId :  
SB-26-9.0-9.5


Tgt Ion:117 Resp: 5084  
Ion Ratio Lower Upper  
117 100  
82 53.2 43.0 64.6  
119 24.9 25.5 38.3#



#72  
1,4-Dichlorobenzene-d4  
Concen: 50.000 ug/l  
RT: 13.422 min Scan# 1903  
Delta R.T. -0.000 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26

Tgt Ion:152 Resp: 1451  
Ion Ratio Lower Upper  
152 100  
115 66.0 28.1 84.2  
150 151.8 0.0 352.0





#76

1,2,3-Trichloropropane

Concen: 55.089 ug/l

RT: 12.483 min Scan# 1749

Delta R.T. -0.146 min

Lab File: VY013406.D

Acq: 24 Apr 2023 17:26

Instrument :

MSVOA\_Y

ClientSampleId :





SB-26-9.0-9.5

Tgt Ion: 75 Resp: 1078

Ion Ratio Lower Upper

75 100

77 0.0 0.0 0.0



#81

trans-1,4-Dichloro-2-butene

Concen: 112.687 ug/l

RT: 12.483 min Scan# 1749

Delta R.T. 0.104 min

Lab File: VY013406.D

Acq: 24 Apr 2023 17:26




Tgt Ion: 75 Resp: 1078


Ion Ratio Lower Upper

75 100

53 0.0 73.2 109.8#

89 0.0 38.0 57.0#









#87  
1,3-Dichlorobenzene  
Concen: 1.697 ug/l  
RT: 13.440 min Scan# 1906  
Delta R.T. 0.073 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26




Instrument :  
MSVOA\_Y  
ClientSampleId :  
SB-26-9.0-9.5


Tgt Ion:146 Resp: 86  
Ion Ratio Lower Upper  
146 100  
111 57.0 19.3 57.9  
148 31.4 31.9 95.7#



#88  
1,4-Dichlorobenzene  
Concen: 1.680 ug/l  
RT: 13.440 min Scan# 1906  
Delta R.T. -0.006 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26


Tgt Ion:146 Resp: 86  
Ion Ratio Lower Upper  
146 100  
111 57.0 18.9 56.5#  
148 31.4 31.9 95.5#





#95  
Naphthalene  
Concen: 2.687 ug/l  
RT: 15.238 min Scan# 21  
Delta R.T. 0.005 min  
Lab File: VY013406.D  
Acq: 24 Apr 2023 17:26

Instrument :  
MSVOA\_Y  
ClientSampleId :  
SB-26-9.0-9.5



Tgt Ion:128 Resp: 159  
Ion Ratio Lower Upper  
128 100  
127 0.0 10.7 16.1#  
129 0.0 8.8 13.2#

