

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042423\
 Data File : VY013413.D
 Acq On : 24 Apr 2023 20:10
 Operator : KP/MD
 Sample : 02371-17
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-16-9.0-9.5

Quant Time: Apr 25 03:18:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	7631	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	12531	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	14137	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	4979	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	8028	95.968	ug/l	0.01
Spiked Amount	50.000	Range	50 - 163	Recovery	=	191.940%#
35) Dibromofluoromethane	7.716	113	6117	77.613	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	155.220%#
50) Toluene-d8	10.179	98	14933	53.261	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	106.520%
62) 4-Bromofluorobenzene	12.477	95	5099	53.946	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	107.900%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042423\
Data File : VY013413.D
Acq On : 24 Apr 2023 20:10
Operator : KP/MD
Sample : 02371-17
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-16-9.0-9.5

Quant Time: Apr 25 03:18:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

