

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042423\
 Data File : VY013430.D
 Acq On : 25 Apr 2023 05:48
 Operator : KP/MD
 Sample : 02423-06
 Misc : 3.94g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 50 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SPD-2WC-23-04-18

Quant Time: Apr 25 07:17:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	178760	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	304366	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	288015	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	132413	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	121719	62.114	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.220%
35) Dibromofluoromethane	7.716	113	103348	53.987	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.980%
50) Toluene-d8	10.179	98	374098	54.933	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	109.860%
62) 4-Bromofluorobenzene	12.477	95	118269	51.515	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	103.020%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042423\
Data File : VY013430.D
Acq On : 25 Apr 2023 05:48
Operator : KP/MD
Sample : 02423-06
Misc : 3.94g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 50 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SPD-2WC-23-04-18

Quant Time: Apr 25 07:17:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

