

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042423\
 Data File : VY013434.D
 Acq On : 25 Apr 2023 07:22
 Operator : KP/MD
 Sample : 02390-06
 Misc : 4.46g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SPC-2WC-23-04-17

Quant Time: Apr 25 08:00:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	165618	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	278083	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	233951	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	79638	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	114981	63.332	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.660%
35) Dibromofluoromethane	7.716	113	97312	55.638	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.280%
50) Toluene-d8	10.179	98	327107	52.573	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	105.140%
62) 4-Bromofluorobenzene	12.483	95	85198	40.617	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	81.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042423\
Data File : VY013434.D
Acq On : 25 Apr 2023 07:22
Operator : KP/MD
Sample : 02390-06
Misc : 4.46g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 54 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SPC-2WC-23-04-17

Quant Time: Apr 25 08:00:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

