

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042423\
 Data File : VY013435.D
 Acq On : 25 Apr 2023 07:44
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 25 08:08:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	89	0.00
2 T	Dichlorodifluoromethane	0.475	0.491	-3.4	90	0.00
3 P	Chloromethane	0.637	0.560	12.1	80	0.00
4 C	Vinyl Chloride	0.634	0.550	13.2#	79	0.00
5 T	Bromomethane	0.506	0.451	10.9	85	0.00
6 T	Chloroethane	0.438	0.355	18.9	77	0.00
7 T	Trichlorofluoromethane	0.961	0.949	1.2	89	0.00
8 T	Diethyl Ether	0.327	0.351	-7.3	94	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.539	0.535	0.7	91	0.00
10 T	Methyl Iodide	0.677	0.683	-0.9	83	0.00
11 T	Tert butyl alcohol	0.066	0.063	4.5	98	0.00
12 CM	1,1-Dichloroethene	0.520	0.519	0.2#	88	0.00
13 T	Acrolein	0.054	0.043	20.4	86	0.00
14 T	Allyl chloride	0.964	0.953	1.1	87	0.00
15 T	Acrylonitrile	0.183	0.200	-9.3	94	0.00
16 T	Acetone	0.216	0.174	19.4	65	0.00
17 T	Carbon Disulfide	1.688	1.721	-2.0	89	0.00
18 T	Methyl Acetate	0.713	0.772	-8.3	95	0.00
19 T	Methyl tert-butyl Ether	1.500	1.608	-7.2	92	0.00
20 T	Methylene Chloride	0.924	0.725	21.5	97	0.00
21 T	trans-1,2-Dichloroethene	0.568	0.592	-4.2	91	0.00
22 T	Diisopropyl ether	1.885	1.970	-4.5	92	0.00
23 T	Vinyl Acetate	1.067	1.092	-2.3	87	0.00
24 P	1,1-Dichloroethane	1.050	1.073	-2.2	91	0.00
25 T	2-Butanone	0.274	0.270	1.5	84	0.00
26 T	2,2-Dichloropropane	0.905	0.814	10.1	80	0.00
27 T	cis-1,2-Dichloroethene	0.628	0.674	-7.3	95	0.00
28 T	Bromochloromethane	0.373	0.377	-1.1	92	0.00
29 T	Tetrahydrofuran	0.170	0.187	-10.0	94	0.00
30 C	Chloroform	1.044	1.117	-7.0#	96	0.00
31 T	Cyclohexane	0.981	0.887	9.6	84	0.00
32 T	1,1,1-Trichloroethane	0.932	0.959	-2.9	91	0.00
33 S	1,2-Dichloroethane-d4	0.548	0.591	-7.8	102	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	92	0.00
35 S	Dibromofluoromethane	0.314	0.327	-4.1	102	0.00
36 T	1,1-Dichloropropene	0.522	0.511	2.1	91	0.00
37 T	Ethyl Acetate	0.393	0.388	1.3	87	0.00
38 T	Carbon Tetrachloride	0.547	0.554	-1.3	92	0.00
39 T	Methylcyclohexane	0.615	0.589	4.2	87	0.00
40 TM	Benzene	1.470	1.509	-2.7	94	0.00
41 T	Methacrylonitrile	0.183	0.165	9.8	81	0.00
42 TM	1,2-Dichloroethane	0.481	0.502	-4.4	94	0.00
43 T	Isopropyl Acetate	0.663	0.685	-3.3	92	0.00
44 TM	Trichloroethene	0.391	0.402	-2.8	94	0.00
45 C	1,2-Dichloropropane	0.384	0.391	-1.8#	93	0.00
46 T	Dibromomethane	0.230	0.245	-6.5	97	0.00
47 T	Bromodichloromethane	0.515	0.546	-6.0	97	0.00
48 T	Methyl methacrylate	0.304	0.320	-5.3	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042423\
 Data File : VY013435.D
 Acq On : 25 Apr 2023 07:44
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 55 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 25 08:08:35 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 02:39:58 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	93	0.00
50 S	Toluene-d8	1.119	1.128	-0.8	99	0.00
51 T	4-Methyl-2-Pentanone	0.369	0.399	-8.1	95	0.00
52 CM	Toluene	0.885	0.924	-4.4#	94	0.00
53 T	t-1,3-Dichloropropene	0.545	0.560	-2.8	92	0.00
54 T	cis-1,3-Dichloropropene	0.613	0.620	-1.1	92	0.00
55 T	1,1,2-Trichloroethane	0.305	0.329	-7.9	98	0.00
56 T	Ethyl methacrylate	0.423	0.473	-11.8	96	0.00
57 T	1,3-Dichloropropane	0.516	0.551	-6.8	97	0.00
58 T	2-Chloroethyl Vinyl ether	0.158	0.160	-1.3	99	0.00
59 T	2-Hexanone	0.269	0.284	-5.6	90	0.00
60 T	Dibromochloromethane	0.370	0.403	-8.9	97	0.00
61 T	1,2-Dibromoethane	0.299	0.322	-7.7	96	0.00
62 S	4-Bromofluorobenzene	0.377	0.394	-4.5	103	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	95	0.00
64 T	Tetrachloroethene	0.380	0.394	-3.7	98	0.00
65 PM	Chlorobenzene	1.089	1.116	-2.5	97	0.00
66 T	1,1,1,2-Tetrachloroethane	0.407	0.414	-1.7	97	0.00
67 C	Ethyl Benzene	1.913	1.958	-2.4#	95	0.00
68 T	m/p-Xylenes	0.727	0.745	-2.5	95	0.00
69 T	o-Xylene	0.688	0.712	-3.5	95	0.00
70 T	Styrene	1.150	1.211	-5.3	97	0.00
71 P	Bromoform	0.285	0.300	-5.3	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	97	0.00
73 T	Isopropylbenzene	3.782	3.776	0.2	94	0.00
74 T	N-amyl acetate	1.346	1.390	-3.3	93	0.00
75 P	1,1,2,2-Tetrachloroethane	0.924	0.960	-3.9	98	0.00
76 T	1,2,3-Trichloropropane	0.674	0.708	-5.0	97	0.00
77 T	Bromobenzene	0.916	0.961	-4.9	99	0.00
78 T	n-propylbenzene	4.658	4.579	1.7	92	0.00
79 T	2-Chlorotoluene	2.601	2.625	-0.9	95	0.00
80 T	1,3,5-Trimethylbenzene	3.087	3.140	-1.7	95	0.00
81 T	trans-1,4-Dichloro-2-butene	0.330	0.310	6.1	85	0.00
82 T	4-Chlorotoluene	2.711	2.688	0.8	94	0.00
83 T	tert-Butylbenzene	2.674	2.743	-2.6	96	0.00
84 T	1,2,4-Trimethylbenzene	3.001	3.087	-2.9	95	0.00
85 T	sec-Butylbenzene	4.014	3.999	0.4	94	0.00
86 T	p-Isopropyltoluene	3.228	3.207	0.7	93	0.00
87 T	1,3-Dichlorobenzene	1.746	1.738	0.5	95	0.00
88 T	1,4-Dichlorobenzene	1.764	1.731	1.9	95	0.00
89 T	n-Butylbenzene	3.150	3.042	3.4	91	0.00
90 T	Hexachloroethane	0.679	0.670	1.3	96	0.00
91 T	1,2-Dichlorobenzene	1.577	1.636	-3.7	98	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.163	0.170	-4.3	100	0.00
93 T	1,2,4-Trichlorobenzene	0.940	0.952	-1.3	97	0.00
94 T	Hexachlorobutadiene	0.547	0.544	0.5	98	0.00
95 T	Naphthalene	2.039	2.340	-14.8	105	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042423\
Data File : VY013435.D
Acq On : 25 Apr 2023 07:44
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 55 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Apr 25 08:08:35 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042423S.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 02:39:58 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.838	0.888	-6.0	100	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6