

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042425\
 Data File : VY021992.D
 Acq On : 24 Apr 2025 11:01
 Operator : SY/MD
 Sample : Q1853-01RE
 Misc : 5.50g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-02-04222025RE

Quant Time: Apr 24 14:11:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	114999	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	170135	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	126145	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	35481	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	44155	41.568	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	83.140%
35) Dibromofluoromethane	7.634	113	59762	53.169	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.340%
50) Toluene-d8	10.109	98	191523	45.197	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	90.400%
62) 4-Bromofluorobenzene	12.408	95	39530	27.711	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	55.420%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042425\
Data File : VY021992.D
Acq On : 24 Apr 2025 11:01
Operator : SY/MD
Sample : Q1853-01RE
Misc : 5.50g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-02-04222025RE

Quant Time: Apr 24 14:11:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

