

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042525\
 Data File : VY022011.D
 Acq On : 25 Apr 2025 16:15
 Operator : SY/MD
 Sample : Q1883-01
 Misc : 9.82g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-PCS-TC-27-042325

Quant Time: Apr 25 22:56:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	302395	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.609	114	575522	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	492784	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	180667	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	136336	48.811	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	97.620%
35) Dibromofluoromethane	7.628	113	185226	48.715	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.440%
50) Toluene-d8	10.103	98	693705	48.395	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.780%
62) 4-Bromofluorobenzene	12.401	95	181125	37.535	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	75.060%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042525\
Data File : VY022011.D
Acq On : 25 Apr 2025 16:15
Operator : SY/MD
Sample : Q1883-01
Misc : 9.82g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-PCS-TC-27-042325

Quant Time: Apr 25 22:56:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

