

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042525\
 Data File : VY022014.D
 Acq On : 25 Apr 2025 17:25
 Operator : SY/MD
 Sample : Q1883-07
 Misc : 11.51g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OU4-PCS-TC-30-042325

Quant Time: Apr 25 22:57:19 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	296986	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	552810	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	505557	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	203067	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.055	65	143279	52.231	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 104.460%	
35) Dibromofluoromethane	7.634	113	186371	51.030	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.060%	
50) Toluene-d8	10.103	98	671074	48.739	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 97.480%	
62) 4-Bromofluorobenzene	12.401	95	196867	42.473	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 84.940%	
Target Compounds						Qvalue
16) Acetone	3.872	43	20052	37.707	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042525\
Data File : VY022014.D
Acq On : 25 Apr 2025 17:25
Operator : SY/MD
Sample : Q1883-07
Misc : 11.51g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OU4-PCS-TC-30-042325

Quant Time: Apr 25 22:57:19 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

