

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042525\
 Data File : VY022019.D
 Acq On : 25 Apr 2025 19:22
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 25 22:58:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	86	0.00
2 T	Dichlorodifluoromethane	0.426	0.384	9.9	84	0.00
3 P	Chloromethane	0.607	0.567	6.6	81	0.00
4 C	Vinyl Chloride	0.745	0.740	0.7#	88	0.00
5 T	Bromomethane	0.642	0.569	11.4	83	0.00
6 T	Chloroethane	0.508	0.506	0.4	89	0.00
7 T	Trichlorofluoromethane	0.983	0.949	3.5	85	0.00
8 T	Diethyl Ether	0.256	0.255	0.4	89	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.533	0.502	5.8	89	0.00
10 T	Methyl Iodide	0.617	0.616	0.2	85	0.00
11 T	Tert butyl alcohol	0.025	0.028	-12.0	93	-0.02
12 CM	1,1-Dichloroethene	0.497	0.473	4.8#	87	0.00
13 T	Acrolein	0.046	0.033	28.3#	67	0.00
14 T	Allyl chloride	0.579	0.571	1.4	89	0.00
15 T	Acrylonitrile	0.094	0.100	-6.4	93	0.00
16 T	Acetone	0.070	0.067	4.3	92	0.00
17 T	Carbon Disulfide	1.585	1.547	2.4	90	0.00
18 T	Methyl Acetate	0.216	0.292	-35.2#	113	0.00
19 T	Methyl tert-butyl Ether	1.113	1.185	-6.5	90	0.00
20 T	Methylene Chloride	0.584	0.547	6.3	89	0.00
21 T	trans-1,2-Dichloroethene	0.557	0.555	0.4	90	0.00
22 T	Diisopropyl ether	1.288	1.374	-6.7	92	0.00
23 T	Vinyl Acetate	0.694	0.715	-3.0	86	0.00
24 P	1,1-Dichloroethane	0.893	0.900	-0.8	91	0.00
25 T	2-Butanone	0.105	0.113	-7.6	92	0.00
26 T	2,2-Dichloropropane	0.782	0.691	11.6	80	0.00
27 T	cis-1,2-Dichloroethene	0.622	0.646	-3.9	91	0.00
28 T	Bromochloromethane	0.347	0.354	-2.0	89	0.00
29 T	Tetrahydrofuran	0.068	0.077	-13.2	91	0.00
30 C	Chloroform	0.990	1.024	-3.4#	93	0.00
31 T	Cyclohexane	0.763	0.717	6.0	87	0.00
32 T	1,1,1-Trichloroethane	0.896	0.887	1.0	90	0.00
33 S	1,2-Dichloroethane-d4	0.462	0.463	-0.2	86	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	88	0.00
35 S	Dibromofluoromethane	0.330	0.325	1.5	87	0.00
36 T	1,1-Dichloropropene	0.444	0.433	2.5	90	0.00
37 T	Ethyl Acetate	0.159	0.167	-5.0	91	0.00
38 T	Carbon Tetrachloride	0.530	0.521	1.7	90	0.00
39 T	Methylcyclohexane	0.558	0.540	3.2	87	0.00
40 TM	Benzene	1.387	1.407	-1.4	93	0.00
41 T	Methacrylonitrile	0.087	0.079	9.2	76	0.00
42 TM	1,2-Dichloroethane	0.343	0.352	-2.6	92	0.00
43 T	Isopropyl Acetate	0.306	0.320	-4.6	90	0.00
44 TM	Trichloroethene	0.384	0.387	-0.8	93	0.00
45 C	1,2-Dichloropropane	0.307	0.315	-2.6#	92	0.00
46 T	Dibromomethane	0.192	0.198	-3.1	93	0.00
47 T	Bromodichloromethane	0.480	0.497	-3.5	93	0.00
48 T	Methyl methacrylate	0.141	0.154	-9.2	90	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042525\
 Data File : VY022019.D
 Acq On : 25 Apr 2025 19:22
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 25 22:58:18 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	97	-0.01
50 S	Toluene-d8	1.245	1.238	0.6	85	0.00
51 T	4-Methyl-2-Pentanone	0.160	0.181	-13.1	93	0.00
52 CM	Toluene	0.898	0.933	-3.9#	92	0.00
53 T	t-1,3-Dichloropropene	0.411	0.420	-2.2	89	0.00
54 T	cis-1,3-Dichloropropene	0.489	0.499	-2.0	91	0.00
55 T	1,1,2-Trichloroethane	0.256	0.275	-7.4	94	0.00
56 T	Ethyl methacrylate	0.293	0.329	-12.3	92	0.00
57 T	1,3-Dichloropropane	0.411	0.435	-5.8	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.143	0.157	-9.8	86	0.00
59 T	2-Hexanone	0.106	0.121	-14.2	93	0.00
60 T	Dibromochloromethane	0.355	0.378	-6.5	94	0.00
61 T	1,2-Dibromoethane	0.242	0.260	-7.4	95	0.00
62 S	4-Bromofluorobenzene	0.419	0.432	-3.1	89	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	92	0.00
64 T	Tetrachloroethene	0.472	0.471	0.2	95	0.00
65 PM	Chlorobenzene	1.118	1.104	1.3	94	0.00
66 T	1,1,1,2-Tetrachloroethane	0.394	0.395	-0.3	94	0.00
67 C	Ethyl Benzene	1.829	1.842	-0.7#	92	0.00
68 T	m/p-Xylenes	0.728	0.744	-2.2	93	0.00
69 T	o-Xylene	0.671	0.699	-4.2	93	0.00
70 T	Styrene	1.126	1.195	-6.1	94	0.00
71 P	Bromoform	0.229	0.242	-5.7	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	96	0.00
73 T	Isopropylbenzene	3.341	3.175	5.0	92	0.00
74 T	N-amyl acetate	0.574	0.556	3.1	91	0.00
75 P	1,1,2,2-Tetrachloroethane	0.563	0.536	4.8	94	0.00
76 T	1,2,3-Trichloropropane	0.428	0.445	-4.0	106	0.00
77 T	Bromobenzene	0.844	0.809	4.1	94	0.00
78 T	n-propylbenzene	3.978	3.772	5.2	91	0.00
79 T	2-Chlorotoluene	2.314	2.199	5.0	93	0.00
80 T	1,3,5-Trimethylbenzene	2.752	2.677	2.7	92	0.00
81 T	trans-1,4-Dichloro-2-butene	0.178	0.154	13.5	84	0.00
82 T	4-Chlorotoluene	2.405	2.274	5.4	92	0.00
83 T	tert-Butylbenzene	2.468	2.394	3.0	93	0.00
84 T	1,2,4-Trimethylbenzene	2.753	2.725	1.0	94	0.00
85 T	sec-Butylbenzene	3.612	3.456	4.3	93	0.00
86 T	p-Isopropyltoluene	3.067	2.943	4.0	92	0.00
87 T	1,3-Dichlorobenzene	1.714	1.626	5.1	95	0.00
88 T	1,4-Dichlorobenzene	1.698	1.605	5.5	96	0.00
89 T	n-Butylbenzene	2.742	2.567	6.4	90	0.00
90 T	Hexachloroethane	0.674	0.620	8.0	94	0.00
91 T	1,2-Dichlorobenzene	1.495	1.455	2.7	97	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.090	0.087	3.3	100	0.00
93 T	1,2,4-Trichlorobenzene	0.847	0.827	2.4	97	0.00
94 T	Hexachlorobutadiene	0.521	0.476	8.6	96	0.00
95 T	Naphthalene	1.408	1.494	-6.1	100	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042525\
Data File : VY022019.D
Acq On : 25 Apr 2025 19:22
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Apr 25 22:58:18 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.731	0.740	-1.2	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6