

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042822\
 Data File : VY008543.D
 Acq On : 28 Apr 2022 13:26
 Operator : KP/MD
 Sample : N2572-01RE
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-14SRE

Quant Time: Apr 29 02:00:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	1183	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	2380	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	3000	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	916	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	1231	88.437	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	176.880%#
35) Dibromofluoromethane	7.716	113	756	49.218	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.440%
50) Toluene-d8	10.185	98	3086	50.397	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.800%
62) 4-Bromofluorobenzene	12.483	95	1057	51.105	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.200%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042822\
Data File : VY008543.D
Acq On : 28 Apr 2022 13:26
Operator : KP/MD
Sample : N2572-01RE
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-14SRE

Quant Time: Apr 29 02:00:18 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 21 14:28:40 2022
Response via : Initial Calibration

