

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042822\
 Data File : VY008551.D
 Acq On : 28 Apr 2022 16:33
 Operator : KP/MD
 Sample : N2600-04
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-15D

Quant Time: Apr 29 02:01:50 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	2522	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	4761	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	4052	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	1408	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	1683	56.715	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 113.440%	
35) Dibromofluoromethane	7.722	113	1259	40.974	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 81.940%	
50) Toluene-d8	10.179	98	5208	42.855	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 85.720%	
62) 4-Bromofluorobenzene	12.489	95	1615	39.033	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 78.060%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042822\
Data File : VY008551.D
Acq On : 28 Apr 2022 16:33
Operator : KP/MD
Sample : N2600-04
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-15D

Quant Time: Apr 29 02:01:50 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 21 14:28:40 2022
Response via : Initial Calibration

