

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042822\
 Data File : VY008552.D
 Acq On : 28 Apr 2022 16:56
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 29 02:02:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	113	0.00
2 T	Dichlorodifluoromethane	0.454	0.446	1.8	111	0.00
3 P	Chloromethane	0.714	0.758	-6.2	114	0.00
4 C	Vinyl Chloride	0.928	1.022	-10.1#	116	0.00
5 T	Bromomethane	0.863	0.919	-6.5	123	0.00
6 T	Chloroethane	0.618	0.678	-9.7	118	0.00
7 T	Trichlorofluoromethane	0.962	0.989	-2.8	116	0.00
8 T	Diethyl Ether	0.331	0.338	-2.1	110	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.573	0.602	-5.1	120	0.00
10 T	Methyl Iodide	0.684	0.691	-1.0	101	0.00
11 T	Tert butyl alcohol	0.068	0.064	5.9	116	0.00
12 CM	1,1-Dichloroethene	0.540	0.550	-1.9#	112	0.00
13 T	Acrolein	0.063	0.049	22.2	87	0.00
14 T	Allyl chloride	0.705	0.751	-6.5	116	0.00
15 T	Acrylonitrile	0.169	0.170	-0.6	107	0.00
16 T	Acetone	0.137	0.119	13.1	89	0.00
17 T	Carbon Disulfide	1.634	1.580	3.3	104	0.00
18 T	Methyl Acetate	0.384	0.368	4.2	108	0.00
19 T	Methyl tert-butyl Ether	1.710	1.782	-4.2	110	0.00
20 T	Methylene Chloride	0.864	0.724	16.2	106	0.00
21 T	trans-1,2-Dichloroethene	0.659	0.677	-2.7	112	0.00
22 T	Diisopropyl ether	1.704	1.835	-7.7	115	0.00
23 T	Vinyl Acetate	1.168	1.275	-9.2	112	0.00
24 P	1,1-Dichloroethane	1.060	1.149	-8.4	117	0.00
25 T	2-Butanone	0.212	0.215	-1.4	105	0.00
26 T	2,2-Dichloropropane	0.982	1.084	-10.4	120	0.00
27 T	cis-1,2-Dichloroethene	0.754	0.823	-9.2	117	0.00
28 T	Bromochloromethane	0.428	0.439	-2.6	105	0.00
29 T	Tetrahydrofuran	0.137	0.143	-4.4	112	0.00
30 C	Chloroform	1.165	1.290	-10.7#	118	0.00
31 T	Cyclohexane	0.967	0.990	-2.4	119	0.00
32 T	1,1,1-Trichloroethane	1.020	1.158	-13.5	123	0.00
33 S	1,2-Dichloroethane-d4	0.588	0.626	-6.5	105	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	112	0.00
35 S	Dibromofluoromethane	0.323	0.349	-8.0	106	0.00
36 T	1,1-Dichloropropene	0.513	0.542	-5.7	117	0.00
37 T	Ethyl Acetate	0.284	0.298	-4.9	117	0.00
38 T	Carbon Tetrachloride	0.533	0.589	-10.5	122	0.00
39 T	Methylcyclohexane	0.657	0.715	-8.8	120	0.00
40 TM	Benzene	1.498	1.625	-8.5	117	0.00
41 T	Methacrylonitrile	0.138	0.138	0.0	125	0.01
42 TM	1,2-Dichloroethane	0.451	0.485	-7.5	114	0.00
43 T	Isopropyl Acetate	0.522	0.565	-8.2	114	0.00
44 TM	Trichloroethene	0.413	0.454	-9.9	122	0.00
45 C	1,2-Dichloropropane	0.360	0.403	-11.9#	118	0.00
46 T	Dibromomethane	0.252	0.270	-7.1	112	0.00
47 T	Bromodichloromethane	0.535	0.599	-12.0	118	0.00
48 T	Methyl methacrylate	0.222	0.244	-9.9	115	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042822\
 Data File : VY008552.D
 Acq On : 28 Apr 2022 16:56
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 29 02:02:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.004	-33.3#	116	0.00
50 S	Toluene-d8	1.176	1.276	-8.5	106	0.00
51 T	4-Methyl-2-Pentanone	0.275	0.301	-9.5	115	0.00
52 CM	Toluene	0.990	1.099	-11.0#	120	0.00
53 T	t-1,3-Dichloropropene	0.574	0.634	-10.5	114	0.00
54 T	cis-1,3-Dichloropropene	0.636	0.687	-8.0	115	0.00
55 T	1,1,2-Trichloroethane	0.344	0.375	-9.0	115	0.00
56 T	Ethyl methacrylate	0.444	0.500	-12.6	115	0.00
57 T	1,3-Dichloropropane	0.561	0.609	-8.6	113	0.00
58 T	2-Chloroethyl Vinyl ether	0.243	0.235	3.3	98	0.00
59 T	2-Hexanone	0.189	0.209	-10.6	113	0.00
60 T	Dibromochloromethane	0.395	0.449	-13.7	117	0.00
61 T	1,2-Dibromoethane	0.340	0.373	-9.7	116	0.00
62 S	4-Bromofluorobenzene	0.435	0.454	-4.4	105	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	109	0.00
64 T	Tetrachloroethene	0.371	0.406	-9.4	125	0.00
65 PM	Chlorobenzene	1.131	1.248	-10.3	120	0.00
66 T	1,1,1,2-Tetrachloroethane	0.417	0.474	-13.7	122	0.00
67 C	Ethyl Benzene	1.962	2.199	-12.1#	123	0.00
68 T	m/p-Xylenes	0.786	0.892	-13.5	124	0.00
69 T	o-Xylene	0.757	0.862	-13.9	124	0.00
70 T	Styrene	1.275	1.452	-13.9	120	0.00
71 P	Bromoform	0.262	0.292	-11.5	118	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	113	0.00
73 T	Isopropylbenzene	3.791	4.224	-11.4	127	0.00
74 T	N-amyl acetate	0.937	1.048	-11.8	118	0.00
75 P	1,1,2,2-Tetrachloroethane	0.847	0.900	-6.3	114	0.00
76 T	1,2,3-Trichloropropane	0.647	0.581	10.2	99	0.00
77 T	Bromobenzene	0.925	0.991	-7.1	119	0.00
78 T	n-propylbenzene	4.547	5.038	-10.8	125	0.00
79 T	2-Chlorotoluene	2.559	2.812	-9.9	123	0.00
80 T	1,3,5-Trimethylbenzene	3.147	3.586	-13.9	130	0.00
81 T	trans-1,4-Dichloro-2-butene	0.279	0.308	-10.4	116	0.00
82 T	4-Chlorotoluene	2.652	2.956	-11.5	123	0.00
83 T	tert-Butylbenzene	2.732	3.128	-14.5	131	0.00
84 T	1,2,4-Trimethylbenzene	3.122	3.617	-15.9	131	0.00
85 T	sec-Butylbenzene	4.122	4.794	-16.3	131	0.00
86 T	p-Isopropyltoluene	3.452	4.034	-16.9	133	0.00
87 T	1,3-Dichlorobenzene	1.851	2.044	-10.4	125	0.00
88 T	1,4-Dichlorobenzene	1.834	2.004	-9.3	125	0.00
89 T	n-Butylbenzene	3.134	3.623	-15.6	128	0.00
90 T	Hexachloroethane	0.621	0.737	-18.7	131	0.00
91 T	1,2-Dichlorobenzene	1.646	1.808	-9.8	125	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.127	0.140	-10.2	120	0.00
93 T	1,2,4-Trichlorobenzene	0.916	1.003	-9.5	122	0.00
94 T	Hexachlorobutadiene	0.447	0.510	-14.1	132	0.00
95 T	Naphthalene	2.077	2.289	-10.2	119	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042822\
Data File : VY008552.D
Acq On : 28 Apr 2022 16:56
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: Apr 29 02:02:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 21 14:28:40 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.802	0.888	-10.7	122	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6