

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042825\
 Data File : VY022030.D
 Acq On : 28 Apr 2025 14:41
 Operator : SY/MD
 Sample : Q1895-03
 Misc : 4.72g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-2

Quant Time: Apr 29 03:54:10 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	316315	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	589559	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	518390	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	195303	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	147955	50.639	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.280%
35) Dibromofluoromethane	7.634	113	195045	50.076	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.160%
50) Toluene-d8	10.109	98	707910	48.210	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.420%
62) 4-Bromofluorobenzene	12.408	95	191573	38.755	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	77.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042825\
Data File : VY022030.D
Acq On : 28 Apr 2025 14:41
Operator : SY/MD
Sample : Q1895-03
Misc : 4.72g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-2

Quant Time: Apr 29 03:54:10 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

