

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042825\
 Data File : VY022021.D
 Acq On : 28 Apr 2025 09:59
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 04/29/2025
 Supervised By : Mahesh Dadoda 04/29/2025

Quant Time: Apr 29 03:51:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	316866	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	463292	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	436582	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	238558	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	133766	45.703	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery =	91.400%		
35) Dibromofluoromethane	7.634	113	150892	49.299	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery =	98.600%		
50) Toluene-d8	10.109	98	576275	49.941	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery =	99.880%		
62) 4-Bromofluorobenzene	12.408	95	195343	50.288	ug/l	0.00
Spiked Amount 50.000	Range 30 - 143		Recovery =	100.580%		
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.867	85	116076	42.991	ug/l	98
3) Chloromethane	2.068	50	173631	45.163	ug/l	99
4) Vinyl Chloride	2.202	62	214876	45.482	ug/l	99
5) Bromomethane	2.592	94	170113	41.794	ug/l	98
6) Chloroethane	2.733	64	146584	45.558	ug/l	100
7) Trichlorofluoromethane	3.056	101	282306	45.316	ug/l	96
8) Diethyl Ether	3.452	74	70779	43.651	ug/l	99
9) 1,1,2-Trichlorotrifluo...	3.812	101	151333	44.780	ug/l	100
10) Methyl Iodide	4.007	142	183769	46.967	ug/l	99
11) Tert butyl alcohol	4.866	59	35986	223.960	ug/l #	93
12) 1,1-Dichloroethene	3.787	96	140007	44.447	ug/l	95
13) Acrolein	3.653	56	39920	138.423	ug/l	96
14) Allyl chloride	4.385	41	163390	44.547	ug/l	99
15) Acrylonitrile	5.061	53	136206	229.455	ug/l	99
16) Acetone	3.873	43	91801	238.658	ug/l	94
17) Carbon Disulfide	4.104	76	452882	45.091	ug/l	99
18) Methyl Acetate	4.391	43	75409	55.156	ug/l	100
19) Methyl tert-butyl Ether	5.116	73	328254	46.559	ug/l	98
20) Methylene Chloride	4.616	84	156196	42.205	ug/l	98
21) trans-1,2-Dichloroethene	5.116	96	162653	46.106	ug/l	99
22) Diisopropyl ether	6.019	45	385976	47.287	ug/l	98
23) Vinyl Acetate	5.964	43	1009046	229.383	ug/l	100
24) 1,1-Dichloroethane	5.915	63	257272	45.472	ug/l #	97
25) 2-Butanone	6.896	43	151911	227.698	ug/l	96
26) 2,2-Dichloropropane	6.884	77	226836	45.744	ug/l	100
27) cis-1,2-Dichloroethene	6.890	96	181037	45.895	ug/l	100
28) Bromochloromethane	7.244	49	102834	46.794	ug/l	99
29) Tetrahydrofuran	7.268	42	105859	244.304	ug/l	100
30) Chloroform	7.421	83	292396	46.592	ug/l	98
31) Cyclohexane	7.701	56	211764	43.769	ug/l	98
32) 1,1,1-Trichloroethane	7.616	97	262924	46.308	ug/l	99
36) 1,1-Dichloropropene	7.835	75	202515	49.257	ug/l	100
37) Ethyl Acetate	6.988	43	72715	49.244	ug/l	98
38) Carbon Tetrachloride	7.817	117	246873	50.250	ug/l	98
39) Methylcyclohexane	9.109	83	248104	47.952	ug/l	97
40) Benzene	8.079	78	631630	49.146	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042825\
 Data File : VY022021.D
 Acq On : 28 Apr 2025 09:59
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050

Manual Integrations
 APPROVED

Reviewed By : Semsettin Yesilyurt 04/29/2025
 Supervised By : Mahesh Dadoda 04/29/2025

Quant Time: Apr 29 03:51:56 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.220	41	40961	50.745	ug/l	96
42) 1,2-Dichloroethane	8.158	62	157468	49.487	ug/l	100
43) Isopropyl Acetate	8.201	43	136500	48.136	ug/l	100
44) Trichloroethene	8.866	130	172375	48.422	ug/l	95
45) 1,2-Dichloropropane	9.140	63	139620	49.124	ug/l	100
46) Dibromomethane	9.231	93	88811	49.917	ug/l	98
47) Bromodichloromethane	9.420	83	222855	50.140	ug/l	100
48) Methyl methacrylate	9.225	41	66889	51.025	ug/l	98
49) 1,4-Dioxane	9.237	88	18782	990.448	ug/l	95
51) 4-Methyl-2-Pentanone	10.000	43	390922	262.978	ug/l	100
52) Toluene	10.170	92	422400	50.774	ug/l	99
53) t-1,3-Dichloropropene	10.396	75	191764	50.300	ug/l	97
54) cis-1,3-Dichloropropene	9.859	75	226727	50.030	ug/l	98
55) 1,1,2-Trichloroethane	10.573	97	121704	51.210	ug/l	98
56) Ethyl methacrylate	10.438	69	138287	50.883	ug/l	99
57) 1,3-Dichloropropane	10.719	76	191690	50.313	ug/l	100
58) 2-Chloroethyl Vinyl ether	9.713	63	364363	274.205	ug/l	99
59) 2-Hexanone	10.762	43	261097	266.294	ug/l	100
60) Dibromochloromethane	10.914	129	169983	51.661	ug/l	100
61) 1,2-Dibromoethane	11.018	107	116441	51.898	ug/l	99
64) Tetrachloroethene	10.646	164	201169	48.844	ug/l	96
65) Chlorobenzene	11.444	112	469987	48.139	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.518	131	171117	49.785	ug/l	99
67) Ethyl Benzene	11.518	91	787306	49.296	ug/l	100
68) m/p-Xylenes	11.627	106	644900	101.477	ug/l	98
69) o-Xylene	11.957	106	294379	50.257	ug/l	99
70) Styrene	11.969	104	508377	51.691	ug/l	99
71) Bromoform	12.133	173	102919	51.502	ug/l #	99
73) Isopropylbenzene	12.255	105	770147	48.321	ug/l	99
74) N-amyl acetate	12.072	43	125014	45.653	ug/l	97
75) 1,1,2,2-Tetrachloroethane	12.505	83	128769	47.921	ug/l	99
76) 1,2,3-Trichloropropane	12.554	75	97818m	47.871	ug/l	
77) Bromobenzene	12.530	156	193416	48.026	ug/l	99
78) n-propylbenzene	12.597	91	918990	48.417	ug/l	99
79) 2-Chlorotoluene	12.682	91	522771	47.359	ug/l	99
80) 1,3,5-Trimethylbenzene	12.737	105	649179	49.434	ug/l	100
81) trans-1,4-Dichloro-2-b...	12.304	75	39219	46.220	ug/l	97
82) 4-Chlorotoluene	12.780	91	552085	48.113	ug/l	100
83) tert-Butylbenzene	12.999	119	564883	47.962	ug/l	99
84) 1,2,4-Trimethylbenzene	13.042	105	652719	49.692	ug/l	99
85) sec-Butylbenzene	13.176	105	833063	48.338	ug/l	99
86) p-Isopropyltoluene	13.292	119	718809	49.118	ug/l	99
87) 1,3-Dichlorobenzene	13.286	146	393855	48.150	ug/l	99
88) 1,4-Dichlorobenzene	13.371	146	383054	47.284	ug/l	99
89) n-Butylbenzene	13.621	91	619718	47.376	ug/l	100
90) Hexachloroethane	13.883	117	149402	46.466	ug/l	99
91) 1,2-Dichlorobenzene	13.657	146	342527	48.011	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.273	75	20568	48.111	ug/l	94
93) 1,2,4-Trichlorobenzene	14.919	180	194321	48.066	ug/l	100
94) Hexachlorobutadiene	15.023	225	119092	47.878	ug/l	98
95) Naphthalene	15.145	128	333693	49.688	ug/l	100
96) 1,2,3-Trichlorobenzene	15.334	180	172423	49.408	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042825\
Data File : VY022021.D
Acq On : 28 Apr 2025 09:59
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050


Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025

Quant Time: Apr 29 03:51:56 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:168 Resp: 316860
Ion Ratio Lower Upper
168 100
99 52.8 43.1 64.7

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025




#2
Dichlorodifluoromethane
Concen: 42.991 ug/l
RT: 1.867 min Scan# 24
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59



Tgt Ion: 85 Resp: 116076
Ion Ratio Lower Upper
85 100
87 32.0 16.6 49.7









#3
Chloromethane
Concen: 45.163 ug/l
RT: 2.068 min Scan# 51
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 50 Resp: 17363
Ion Ratio Lower Upper
50 100
52 32.1 26.1 39.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025



#4
Vinyl Chloride
Concen: 45.482 ug/l
RT: 2.202 min Scan# 79
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59


Tgt Ion: 62 Resp: 214876
Ion Ratio Lower Upper
62 100
64 31.9 25.1 37.7





#5
Bromomethane
Concen: 41.794 ug/l
RT: 2.592 min Scan# 143
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 94 Resp: 17011
Ion Ratio Lower Upper
94 100
96 95.2 74.6 111.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025




#6
Chloroethane
Concen: 45.558 ug/l
RT: 2.733 min Scan# 166
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59



Tgt Ion: 64 Resp: 146584
Ion Ratio Lower Upper
64 100
66 32.8 26.2 39.4









#7
Trichlorofluoromethane
Concen: 45.316 ug/l
RT: 3.056 min Scan# 218
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:101 Resp: 282300
Ion Ratio Lower Upper
101 100
103 63.6 53.6 80.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025



#8
Diethyl Ether
Concen: 43.651 ug/l
RT: 3.452 min Scan# 284
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59


Tgt Ion: 74 Resp: 70779
Ion Ratio Lower Upper
74 100
45 79.4 39.1 117.5





#9
1,1,2-Trichlorotrifluoroethane
Concen: 44.780 ug/l
RT: 3.812 min Scan# 343
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:101 Resp: 15133
Ion Ratio Lower Upper
101 100
85 43.8 34.6 51.8
151 79.7 63.7 95.5

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025




#10
Methyl Iodide
Concen: 46.967 ug/l
RT: 4.007 min Scan# 375
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59



Tgt Ion:142 Resp: 183769
Ion Ratio Lower Upper
142 100
127 46.0 36.2 54.2
141 14.9 11.7 17.5





#11

Tert butyl alcohol

Concen: 223.960 ug/l

RT: 4.866 min Scan# 518

Delta R.T. -0.012 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 59 Resp: 35980

Ion Ratio Lower Upper

59 100





57 9.0 9.3 13.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/29/2025

Supervised By :Mahesh Dadoda 04/29/2025



#12

1,1-Dichloroethene

Concen: 44.447 ug/l

RT: 3.787 min Scan# 339

Delta R.T. -0.000 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59




Tgt Ion: 96 Resp: 140007


Ion Ratio Lower Upper

96 100

61 147.6 113.0 169.4


98 69.0 51.4 77.2





#13
Acrolein
Concen: 138.423 ug/l
RT: 3.653 min Scan# 317
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 56 Resp: 39920
Ion Ratio Lower Upper
56 100
55 75.3 57.4 86.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025




#14
Allyl chloride
Concen: 44.547 ug/l
RT: 4.385 min Scan# 437
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Tgt Ion: 41 Resp: 163390
Ion Ratio Lower Upper
41 100
39 70.5 57.2 85.8
76 48.5 39.7 59.5





#15
Acrylonitrile
Concen: 229.455 ug/l
RT: 5.061 min Scan# 54
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 53 Resp: 136200
Ion Ratio Lower Upper
53 100
52 81.4 65.0 97.6
51 37.9 28.9 43.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025



#16
Acetone
Concen: 238.658 ug/l
RT: 3.873 min Scan# 353
Delta R.T. -0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59


Tgt Ion: 43 Resp: 91801
Ion Ratio Lower Upper
43 100
58 37.2 27.0 40.4





#17
Carbon Disulfide
Concen: 45.091 ug/l
RT: 4.104 min Scan# 391
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 76 Resp: 45288
Ion Ratio Lower Upper
76 100
78 9.1 7.6 11.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025



#18
Methyl Acetate
Concen: 55.156 ug/l
RT: 4.391 min Scan# 438
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59


Tgt Ion: 43 Resp: 75409
Ion Ratio Lower Upper
43 100
74 30.5 24.2 36.4





#19
Methyl tert-butyl Ether
Concen: 46.559 ug/l
RT: 5.116 min Scan# 51
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 73 Resp: 328254
Ion Ratio Lower Upper
73 100
57 20.3 16.9 25.3

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025




#20
Methylene Chloride
Concen: 42.205 ug/l
RT: 4.616 min Scan# 475
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Tgt Ion: 84 Resp: 156196
Ion Ratio Lower Upper
84 100
49 99.3 79.9 119.9
51 31.2 25.4 38.0
86 59.4 50.6 76.0





#21
trans-1,2-Dichloroethene
Concen: 46.106 ug/l
RT: 5.116 min Scan# 511
Delta R.T. 0.012 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 96 Resp: 16265
Ion Ratio Lower Upper
96 100
61 116.2 92.6 138.8
98 62.1 50.2 75.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025



#22
Diisopropyl ether
Concen: 47.287 ug/l
RT: 6.019 min Scan# 705
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59


Tgt Ion: 45 Resp: 385976
Ion Ratio Lower Upper
45 100
43 56.2 43.6 65.4
87 34.1 28.2 42.4
59 12.9 10.0 15.0





#23
 Vinyl Acetate
 Concen: 229.383 ug/l
 RT: 5.964 min Scan# 695
 Delta R.T. 0.006 min
 Lab File: VY022021.D
 Acq: 28 Apr 2025 09:59




Instrument :
 MSVOA_Y
 ClientSampleId :
 VSTDCCC050




Tgt Ion: 43 Resp: 100904
 Ion Ratio Lower Upper
 43 100
 86 13.8 11.1 16.7

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 04/29/2025
 Supervised By :Mahesh Dadoda 04/29/2025




#24
 1,1-Dichloroethane
 Concen: 45.472 ug/l
 RT: 5.915 min Scan# 688
 Delta R.T. -0.000 min
 Lab File: VY022021.D
 Acq: 28 Apr 2025 09:59



Tgt Ion: 63 Resp: 257272
 Ion Ratio Lower Upper
 63 100
 98 8.1 4.0 11.9
 100 6.1 2.0 6.0#









#25
2-Butanone
Concen: 227.698 ug/l
RT: 6.896 min Scan# 849
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 15191
Ion Ratio Lower Upper
43 100
72 36.2 27.0 40.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025



#26
2,2-Dichloropropane
Concen: 45.744 ug/l
RT: 6.884 min Scan# 847
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59

Tgt Ion: 77 Resp: 226836
Ion Ratio Lower Upper
77 100
97 24.9 12.4 37.4





#27

cis-1,2-Dichloroethene

Concen: 45.895 ug/l

RT: 6.890 min Scan# 848

Delta R.T. -0.000 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 96 Resp: 18103

Ion Ratio Lower Upper

96 100

61 125.4 0.0 251.4





98 64.2 0.0 128.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/29/2025

Supervised By :Mahesh Dadoda 04/29/2025



#28

Bromochloromethane

Concen: 46.794 ug/l

RT: 7.244 min Scan# 906

Delta R.T. -0.000 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59




Tgt Ion: 49 Resp: 102834


Ion Ratio Lower Upper

49 100

129 3.0 0.0 5.0


130 99.2 78.2 117.4





#29
Tetrahydrofuran
Concen: 244.304 ug/l
RT: 7.268 min Scan# 910
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 42 Resp: 105859
Ion Ratio Lower Upper
42 100
72 53.4 42.6 63.8
71 50.7 40.6 61.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025




#30
Chloroform
Concen: 46.592 ug/l
RT: 7.421 min Scan# 935
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Tgt Ion: 83 Resp: 292396
Ion Ratio Lower Upper
83 100
85 62.1 51.3 76.9





#31
Cyclohexane
Concen: 43.769 ug/l
RT: 7.701 min Scan# 981
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 56 Resp: 211764
Ion Ratio Lower Upper
56 100
69 36.9 29.4 44.2
84 105.9 82.2 123.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025




#32
1,1,1-Trichloroethane
Concen: 46.308 ug/l
RT: 7.616 min Scan# 967
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59



Tgt Ion: 97 Resp: 262924
Ion Ratio Lower Upper
97 100
99 64.4 51.4 77.2
61 38.4 31.4 47.2





#33

1,2-Dichloroethane-d4

Concen: 45.703 ug/l

RT: 8.061 min Scan# 1040

Delta R.T. -0.000 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 65 Resp: 133760

Ion Ratio Lower Upper

65 100





67 53.2 0.0 105.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/29/2025

Supervised By :Mahesh Dadoda 04/29/2025



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. -0.000 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59




Tgt Ion:114 Resp: 463292


Ion Ratio Lower Upper

114 100

63 17.1 0.0 35.4

88 14.3 0.0 28.2









#35
Dibromofluoromethane
Concen: 49.299 ug/l
RT: 7.634 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:113 Resp: 150892
Ion Ratio Lower Upper
113 100
111 101.7 81.8 122.8
192 20.5 15.6 23.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025



#36
1,1-Dichloropropene
Concen: 49.257 ug/l
RT: 7.835 min Scan# 1003
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59

Tgt Ion: 75 Resp: 202515
Ion Ratio Lower Upper
75 100
110 38.3 19.1 57.3
77 31.8 25.5 38.3









#37
Ethyl Acetate
Concen: 49.244 ug/l
RT: 6.988 min Scan# 80
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 43 Resp: 72715
Ion Ratio Lower Upper
43 100
61 16.6 13.8 20.6
70 13.1 9.2 13.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025



#38
Carbon Tetrachloride
Concen: 50.250 ug/l
RT: 7.817 min Scan# 1000
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59

Tgt Ion:117 Resp: 246873
Ion Ratio Lower Upper
117 100
119 97.4 76.2 114.4
121 29.9 24.0 36.0





#39

Methylcyclohexane

Concen: 47.952 ug/l

RT: 9.109 min Scan# 1

Delta R.T. -0.000 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion: 83 Resp: 248104

Ion	Ratio	Lower	Upper
83	100		
55	66.0	50.6	75.8
98	50.3	38.6	58.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025



#40

Benzene

Concen: 49.146 ug/l

RT: 8.079 min Scan# 1043




Delta R.T. -0.000 min


Lab File: VY022021.D

Acq: 28 Apr 2025 09:59

Tgt Ion: 78 Resp: 631630


Ion	Ratio	Lower	Upper
78	100		
77	24.7	19.3	28.9





#41
Methacrylonitrile
Concen: 50.745 ug/l
RT: 7.220 min Scan# 903
Delta R.T. -0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 41 Resp: 4096
Ion Ratio Lower Upper
41 100
39 51.7 36.6 55.0
67 92.5 72.2 108.4
52 35.7 27.6 41.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025



#42
1,2-Dichloroethane
Concen: 49.487 ug/l
RT: 8.158 min Scan# 1056
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59


Tgt Ion: 62 Resp: 157468
Ion Ratio Lower Upper
62 100
98 11.0 0.0 22.0





#43
Isopropyl Acetate
Concen: 48.136 ug/l
RT: 8.201 min Scan# 1063
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 136500
Ion Ratio Lower Upper
43 100
61 23.9 18.9 28.3
87 17.0 13.7 20.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025



#44
Trichloroethene
Concen: 48.422 ug/l
RT: 8.866 min Scan# 1172
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59

Tgt Ion:130 Resp: 172375
Ion Ratio Lower Upper
130 100
95 96.5 0.0 183.2





#45

1,2-Dichloropropane

Concen: 49.124 ug/l

RT: 9.140 min Scan# 1217

Delta R.T. -0.000 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 63 Resp: 139620

Ion Ratio Lower Upper

63 100





65 31.3 25.2 37.8

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/29/2025

Supervised By :Mahesh Dadoda 04/29/2025



#46

Dibromomethane

Concen: 49.917 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. -0.000 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59




Tgt Ion: 93 Resp: 88811


Ion Ratio Lower Upper

93 100

95 83.3 66.7 100.1


174 108.5 84.6 127.0





#47
Bromodichloromethane
Concen: 50.140 ug/l
RT: 9.420 min Scan# 1263
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 83 Resp: 22285
Ion Ratio Lower Upper
83 100
85 64.8 51.8 77.8
127 8.7 7.0 10.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025




#48
Methyl methacrylate
Concen: 51.025 ug/l
RT: 9.225 min Scan# 1231
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59



Tgt Ion: 41 Resp: 66889
Ion Ratio Lower Upper
41 100
69 109.8 85.5 128.3
39 57.9 46.2 69.4





#49

1,4-Dioxane

Concen: 990.448 ug/l

RT: 9.237 min Scan# 1233

Delta R.T. -0.000 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050

Tgt Ion:	88	Resp:	1878
Ion	Ratio	Lower	Upper
88	100		
43	24.2	21.3	31.9
58	59.0	50.5	75.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025



#50

Toluene-d8

Concen: 49.941 ug/l




RT: 10.109 min Scan# 1376


Delta R.T. 0.006 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59


Tgt Ion:	98	Resp:	576275
Ion	Ratio	Lower	Upper
98	100		
100	65.1	51.6	77.4





#51
4-Methyl-2-Pentanone
Concen: 262.978 ug/l
RT: 10.000 min Scan# 1358
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 43 Resp: 390922
Ion Ratio Lower Upper
43 100
58 45.6 36.6 55.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025




#52
Toluene
Concen: 50.774 ug/l
RT: 10.170 min Scan# 1386
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Tgt Ion: 92 Resp: 422400
Ion Ratio Lower Upper
92 100
91 171.4 138.6 207.8





#53
t-1,3-Dichloropropene
Concen: 50.300 ug/l
RT: 10.396 min Scan# 1422
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 75 Resp: 191764
Ion Ratio Lower Upper
75 100
77 31.4 24.0 36.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025




#54
cis-1,3-Dichloropropene
Concen: 50.030 ug/l
RT: 9.859 min Scan# 1335
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Tgt Ion: 75 Resp: 226727
Ion Ratio Lower Upper
75 100
77 31.7 26.6 40.0
39 36.0 29.6 44.4





#55
1,1,2-Trichloroethane
Concen: 51.210 ug/l
RT: 10.573 min Scan# 1452
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59

Instrument : MSVOA_Y
ClientSampleId : VSTDCCC050






Tgt Ion: 97 Resp: 121704


Ion	Ratio	Lower	Upper
97	100		
83	84.8	66.7	100.1
85	54.4	42.6	64.0
99	64.6	49.7	74.5

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025






#56
Ethyl methacrylate
Concen: 50.883 ug/l
RT: 10.438 min Scan# 1430
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59



Tgt Ion: 69 Resp: 138287


Ion	Ratio	Lower	Upper
69	100		
41	51.3	40.7	61.1
39	28.0	23.3	34.9





#57
1,3-Dichloropropane
Concen: 50.313 ug/l
RT: 10.719 min Scan# 1475
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 76 Resp: 191690
Ion Ratio Lower Upper
76 100
78 33.4 26.6 39.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025



#58
2-Chloroethyl Vinyl ether
Concen: 274.205 ug/l
RT: 9.713 min Scan# 1311
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59


Tgt Ion: 63 Resp: 364363
Ion Ratio Lower Upper
63 100
106 32.3 25.3 37.9





#59
2-Hexanone
Concen: 266.294 ug/l
RT: 10.762 min Scan# 1483
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 43 Resp: 261097
Ion Ratio Lower Upper
43 100
58 63.9 32.0 96.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025



#60
Dibromochloromethane
Concen: 51.661 ug/l
RT: 10.914 min Scan# 1508
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59


Tgt Ion:129 Resp: 169983
Ion Ratio Lower Upper
129 100
127 76.9 38.5 115.3





#61
1,2-Dibromoethane
Concen: 51.898 ug/l
RT: 11.018 min Scan# 1524
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:107 Resp: 11644
Ion Ratio Lower Upper
107 100
109 93.7 75.8 113.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025




#62
4-Bromofluorobenzene
Concen: 50.288 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Tgt Ion: 95 Resp: 195343
Ion Ratio Lower Upper
95 100
174 89.9 0.0 179.0
176 87.5 0.0 171.8





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.420 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:117 Resp: 43658
Ion Ratio Lower Upper
117 100
82 50.5 42.1 63.1
119 32.4 25.7 38.5

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025




#64
Tetrachloroethene
Concen: 48.844 ug/l
RT: 10.646 min Scan# 1464
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Tgt Ion:164 Resp: 201169
Ion Ratio Lower Upper
164 100
166 123.8 103.6 155.4
129 91.2 74.2 111.4
131 87.0 72.6 108.8





#65
Chlorobenzene
Concen: 48.139 ug/l
RT: 11.444 min Scan# 11
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion:112 Resp: 46998
Ion Ratio Lower Upper
112 100
114 32.0 25.0 37.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025



#66
1,1,1,2-Tetrachloroethane
Concen: 49.785 ug/l
RT: 11.518 min Scan# 1607
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59

Tgt Ion:131 Resp: 171117
Ion Ratio Lower Upper
131 100
133 94.2 47.8 143.3
119 65.1 32.4 97.0





#67

Ethyl Benzene

Concen: 49.296 ug/l

RT: 11.518 min Scan# 1607

Delta R.T. -0.000 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion: 91 Resp: 787300

Ion Ratio Lower Upper

91 100





106 32.6 25.9 38.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/29/2025

Supervised By :Mahesh Dadoda 04/29/2025



#68

m/p-Xylenes

Concen: 101.477 ug/l

RT: 11.627 min Scan# 1625

Delta R.T. -0.000 min

Lab File: VY022021.D




Acq: 28 Apr 2025 09:59


Tgt Ion:106 Resp: 644900

Ion Ratio Lower Upper

106 100

91 192.6 156.2 234.2









#69
o-Xylene
Concen: 50.257 ug/l
RT: 11.957 min Scan# 1679
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59

Instrument : MSVOA_Y
ClientSampleId : VSTDCCC050

Tgt Ion:106 Resp: 294379

Ion	Ratio	Lower	Upper
106	100		
91	204.8	103.0	308.9




Manual Integrations
APPROVED
Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025




#70
Styrene
Concen: 51.691 ug/l
RT: 11.969 min Scan# 1681
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59

Tgt Ion:104 Resp: 508377

Ion	Ratio	Lower	Upper
104	100		
78	45.4	36.6	54.8
103	54.1	43.8	65.8





#71

Bromoform

Concen: 51.502 ug/l

RT: 12.133 min Scan# 1708

Delta R.T. -0.000 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:173 Resp: 102919

Ion Ratio Lower Upper

173 100

175 48.1 24.3 72.8





254 0.1 0.0 0.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/29/2025

Supervised By :Mahesh Dadoda 04/29/2025



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.346 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59




Tgt Ion:152 Resp: 238558


Ion Ratio Lower Upper

152 100

115 55.8 28.0 84.0

150 172.0 0.0 345.6





#73

Isopropylbenzene

Concen: 48.321 ug/l

RT: 12.255 min Scan# 1

Delta R.T. -0.000 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:105 Resp: 77014

Ion Ratio Lower Upper

105 100





120 27.3 13.5 40.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/29/2025

Supervised By :Mahesh Dadoda 04/29/2025



#74

N-ethyl acetate

Concen: 45.653 ug/l

RT: 12.072 min Scan# 1698

Delta R.T. 0.006 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59

Tgt Ion: 43 Resp: 125014




Ion Ratio Lower Upper


43 100

70 56.2 43.2 64.8

55 34.5 25.7 38.5


61 28.7 23.0 34.6





#75
1,1,2,2-Tetrachloroethane
Concen: 47.921 ug/l
RT: 12.505 min Scan# 1769
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 83 Resp: 128769
Ion Ratio Lower Upper
83 100
131 11.3 5.5 16.4
85 65.4 32.3 96.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025




#76
1,2,3-Trichloropropane
Concen: 47.871 ug/l m
RT: 12.554 min Scan# 1777
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59



Tgt Ion: 75 Resp: 97818
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 48.026 ug/l

RT: 12.530 min Scan# 1773

Delta R.T. -0.000 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050

Tgt Ion:	156	Resp:	193410
Ion Ratio	Lower	Upper	
156	100		
77	164.1	82.7	248.1
158	97.1	48.1	144.4

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025



#78

n-propylbenzene

Concen: 48.417 ug/l




RT: 12.597 min Scan# 1784


Delta R.T. 0.006 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59


Tgt Ion:	91	Resp:	918990
Ion Ratio	Lower	Upper	
91	100		
120	24.1	11.9	35.5





#79
2-Chlorotoluene
Concen: 47.359 ug/l
RT: 12.682 min Scan# 1797
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion: 91 Resp: 52277
Ion Ratio Lower Upper
91 100
126 36.4 17.9 53.8

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025




#80
1,3,5-Trimethylbenzene
Concen: 49.434 ug/l
RT: 12.737 min Scan# 1807
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59



Tgt Ion:105 Resp: 649179
Ion Ratio Lower Upper
105 100
120 50.3 25.0 75.0









#81
trans-1,4-Dichloro-2-butene
Concen: 46.220 ug/l
RT: 12.304 min Scan# 1735
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion: 75 Resp: 39219
Ion Ratio Lower Upper
75 100
53 85.3 66.0 99.0
89 48.4 37.2 55.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025



#82
4-Chlorotoluene
Concen: 48.113 ug/l
RT: 12.780 min Scan# 1814
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59


Tgt Ion: 91 Resp: 552085
Ion Ratio Lower Upper
91 100
126 36.0 17.9 53.8





#83
tert-Butylbenzene
Concen: 47.962 ug/l
RT: 12.999 min Scan# 1850
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:119 Resp: 56488
Ion Ratio Lower Upper
119 100
91 61.1 30.3 90.9
134 26.8 12.9 38.7

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025




#84
1,2,4-Trimethylbenzene
Concen: 49.692 ug/l
RT: 13.042 min Scan# 1857
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59



Tgt Ion:105 Resp: 652719
Ion Ratio Lower Upper
105 100
120 45.9 23.4 70.0





#85

sec-Butylbenzene

Concen: 48.338 ug/l

RT: 13.176 min Scan# 1879

Delta R.T. -0.000 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050

Tgt Ion:105 Resp: 833063

Ion Ratio Lower Upper

105 100





134 20.8 10.3 30.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/29/2025

Supervised By :Mahesh Dadoda 04/29/2025



#86

p-Isopropyltoluene

Concen: 49.118 ug/l

RT: 13.292 min Scan# 1898

Delta R.T. -0.000 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59




Tgt Ion:119 Resp: 718809


Ion Ratio Lower Upper

119 100

134 26.7 13.6 40.7


91 22.2 11.0 33.0





#87
1,3-Dichlorobenzene
Concen: 48.150 ug/l
RT: 13.286 min Scan# 1897
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050




Tgt Ion:146 Resp: 39385
Ion Ratio Lower Upper
146 100
111 37.8 19.1 57.5
148 63.1 32.0 96.2

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025




#88
1,4-Dichlorobenzene
Concen: 47.284 ug/l
RT: 13.371 min Scan# 1911
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Tgt Ion:146 Resp: 383054
Ion Ratio Lower Upper
146 100
111 37.7 19.2 57.6
148 63.4 32.0 96.0





#89
n-Butylbenzene
Concen: 47.376 ug/l
RT: 13.621 min Scan# 1952
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59




Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050



Tgt Ion: 91 Resp: 619718
Ion Ratio Lower Upper
91 100
92 53.3 26.5 79.5
134 26.8 13.4 40.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025



#90
Hexachloroethane
Concen: 46.466 ug/l
RT: 13.883 min Scan# 1995
Delta R.T. 0.006 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59

Tgt Ion:117 Resp: 149402
Ion Ratio Lower Upper
117 100
201 78.4 38.7 116.1









#91
1,2-Dichlorobenzene
Concen: 48.011 ug/l
RT: 13.657 min Scan# 1958
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59

Instrument :
MSVOA_Y
ClientSampleId :
VSTDCCC050

Tgt Ion:146 Resp: 34252
Ion Ratio Lower Upper
146 100
111 39.8 19.7 59.1
148 64.5 32.1 96.3




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025



#92
1,2-Dibromo-3-Chloropropane
Concen: 48.111 ug/l
RT: 14.273 min Scan# 2059
Delta R.T. -0.000 min
Lab File: VY022021.D
Acq: 28 Apr 2025 09:59

Tgt Ion: 75 Resp: 20568
Ion Ratio Lower Upper
75 100
155 95.8 50.7 152.3
157 117.7 62.6 187.8





#93

1,2,4-Trichlorobenzene

Concen: 48.066 ug/l

RT: 14.919 min Scan# 2165

Delta R.T. -0.000 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59

Instrument :

MSVOA_Y

ClientSampleId :





VSTDCCC050

Tgt Ion:180	Resp: 19432
Ion Ratio	Lower Upper
180 100	
182 94.5	47.1 141.4
145 30.2	15.0 45.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025



#94

Hexachlorobutadiene

Concen: 47.878 ug/l




RT: 15.023 min Scan# 2182


Delta R.T. -0.000 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59

Tgt Ion:225	Resp: 119092
Ion Ratio	Lower Upper
225 100	
223 61.9	31.9 95.5
227 63.6	32.3 96.9





#95

Naphthalene

Concen: 49.688 ug/l

RT: 15.145 min Scan# 2201

Delta R.T. 0.006 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59

Instrument :

MSVOA_Y

ClientSampleId :

VSTDCCC050





Tgt Ion:128 Resp: 33369

Ion	Ratio	Lower	Upper
128	100		
127	13.2	10.6	16.0
129	11.2	8.9	13.3

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 04/29/2025
Supervised By :Mahesh Dadoda 04/29/2025



#96

1,2,3-Trichlorobenzene

Concen: 49.408 ug/l

RT: 15.334 min Scan# 2233

Delta R.T. 0.006 min

Lab File: VY022021.D

Acq: 28 Apr 2025 09:59

Tgt Ion:180 Resp: 172423

Ion	Ratio	Lower	Upper
180	100		
182	94.7	48.1	144.4
145	31.3	16.0	48.0

