

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042922\
 Data File : VY008563.D
 Acq On : 29 Apr 2022 16:47
 Operator : KP/MD
 Sample : N2600-04
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-15D

Quant Time: Apr 30 00:38:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	32423	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	57250	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	89433	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	23225	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	17615	46.173	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	92.340%
35) Dibromofluoromethane	7.716	113	18341	49.640	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.280%
50) Toluene-d8	10.179	98	109051	73.019	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	146.040%#
62) 4-Bromofluorobenzene	12.483	95	21314	42.840	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	85.680%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.722	84	8485	13.907	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042922\
Data File : VY008563.D
Acq On : 29 Apr 2022 16:47
Operator : KP/MD
Sample : N2600-04
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-15D

Quant Time: Apr 30 00:38:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 21 14:28:40 2022
Response via : Initial Calibration

