

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042922\
 Data File : VY008565.D
 Acq On : 29 Apr 2022 17:33
 Operator : KP/MD
 Sample : N2599-01RE
 Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 M-1RE

Quant Time: Apr 30 00:39:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
 Quant Title : SW846 8260
 QLast Update : Thu Apr 21 14:28:40 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	8468	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	22446	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	11310	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	4143	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	5898	59.195	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	118.400%	
35) Dibromofluoromethane	7.716	113	4133	28.531	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	57.060%	
50) Toluene-d8	10.179	98	25282	44.063	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	88.120%	
62) 4-Bromofluorobenzene	12.483	95	4361	22.357	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	44.720%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.716	84	1993	11.798	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042922\
Data File : VY008565.D
Acq On : 29 Apr 2022 17:33
Operator : KP/MD
Sample : N2599-01RE
Misc : 5.07g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
M-1RE

Quant Time: Apr 30 00:39:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042122S.M
Quant Title : SW846 8260
QLast Update : Thu Apr 21 14:28:40 2022
Response via : Initial Calibration

