

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042925\
 Data File : VY022050.D
 Acq On : 29 Apr 2025 11:56
 Operator : SY/MD
 Sample : IBLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 IBLK

Quant Time: Apr 30 01:33:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	337580	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	620881	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	526854	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	184068	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	144932	46.480	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	92.960%
35) Dibromofluoromethane	7.634	113	201379	49.094	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	98.180%
50) Toluene-d8	10.109	98	724399	46.844	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.680%
62) 4-Bromofluorobenzene	12.408	95	190299	36.555	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	73.120%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042925\
Data File : VY022050.D
Acq On : 29 Apr 2025 11:56
Operator : SY/MD
Sample : IBLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
IBLK

Quant Time: Apr 30 01:33:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

