

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042925\
 Data File : VY022053.D
 Acq On : 29 Apr 2025 13:55
 Operator : SY/MD
 Sample : Q1892-03
 Misc : 6.04g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-G-VOC

Quant Time: Apr 30 01:33:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	318450	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	610705	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	534852	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	200266	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	156935	53.353	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 106.700%	
35) Dibromofluoromethane	7.634	113	203599	50.463	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 100.920%	
50) Toluene-d8	10.109	98	723434	47.561	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 95.120%	
62) 4-Bromofluorobenzene	12.408	95	199837	39.027	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 78.060%	
Target Compounds						
96) 1,2,3-Trichlorobenzene	15.328	180	3154	1.077	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042925\
Data File : VY022053.D
Acq On : 29 Apr 2025 13:55
Operator : SY/MD
Sample : Q1892-03
Misc : 6.04g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-G-VOC

Quant Time: Apr 30 01:33:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

