

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042925\
 Data File : VY022058.D
 Acq On : 29 Apr 2025 15:53
 Operator : SY/MD
 Sample : Q1906-11
 Misc : 5.23g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-6-VOC

Quant Time: Apr 30 01:34:41 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	299098	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	573151	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	510156	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	191646	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	148128	53.617	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	107.240%
35) Dibromofluoromethane	7.634	113	194414	51.343	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.680%
50) Toluene-d8	10.103	98	692276	48.495	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.980%
62) 4-Bromofluorobenzene	12.408	95	189524	39.438	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	78.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042925\
Data File : VY022058.D
Acq On : 29 Apr 2025 15:53
Operator : SY/MD
Sample : Q1906-11
Misc : 5.23g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-6-VOC

Quant Time: Apr 30 01:34:41 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

