

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042925\
 Data File : VY022062.D
 Acq On : 29 Apr 2025 17:27
 Operator : SY/MD
 Sample : Q1903-02
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 COMP-5

Quant Time: Apr 30 01:35:23 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	293981	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	545754	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	490384	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	184808	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	139187	51.257	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 102.520%	
35) Dibromofluoromethane	7.634	113	184342	51.127	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.260%	
50) Toluene-d8	10.103	98	667552	49.110	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 98.220%	
62) 4-Bromofluorobenzene	12.408	95	183504	40.102	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 80.200%	
Target Compounds						
16) Acetone	3.879	43	10855	10.038	ug/l	Qvalue 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY042925\
Data File : VY022062.D
Acq On : 29 Apr 2025 17:27
Operator : SY/MD
Sample : Q1903-02
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
COMP-5

Quant Time: Apr 30 01:35:23 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

