

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050124\
 Data File : VY018101.D
 Acq On : 01 May 2024 16:39
 Operator : SY/MD
 Sample : P2310-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SU-02-042924

Quant Time: May 02 05:50:21 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050124S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 16:21:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	309720	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	543390	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	474319	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	190795	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	123774	50.699	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	101.400%
35) Dibromofluoromethane	7.710	113	158198	50.661	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.320%
50) Toluene-d8	10.179	98	593241	48.663	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.320%
62) 4-Bromofluorobenzene	12.477	95	193013	46.006	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.020%
Target Compounds						
20) Methylene Chloride	4.704	84	21360	6.554	ug/l	Qvalue 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050124\
Data File : VY018101.D
Acq On : 01 May 2024 16:39
Operator : SY/MD
Sample : P2310-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SU-02-042924

Quant Time: May 02 05:50:21 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050124S.M
Quant Title : SW846 8260
QLast Update : Wed May 01 16:21:20 2024
Response via : Initial Calibration

