

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
 Data File : VY022104.D
 Acq On : 01 May 2025 13:51
 Operator : SY/MD
 Sample : Q1922-07
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-2-VOC

Quant Time: May 02 01:30:06 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	305268	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	575312	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	506620	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	187651	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	144665	51.305	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.620%
35) Dibromofluoromethane	7.634	113	191904	50.490	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.980%
50) Toluene-d8	10.109	98	692091	48.300	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	96.600%
62) 4-Bromofluorobenzene	12.408	95	184357	38.219	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	76.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050125\
Data File : VY022104.D
Acq On : 01 May 2025 13:51
Operator : SY/MD
Sample : Q1922-07
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-2-VOC

Quant Time: May 02 01:30:06 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

