

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050322\
 Data File : VY008588.D
 Acq On : 03 May 2022 13:45
 Operator : KP/MD
 Sample : N2605-01
 Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-04-042822

Quant Time: May 03 16:57:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 01:57:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	51012	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	77910	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	65181	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	26724	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	16322	32.921	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	65.840%
35) Dibromofluoromethane	7.716	113	22714	44.060	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	88.120%
50) Toluene-d8	10.179	98	88413	48.408	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	96.820%
62) 4-Bromofluorobenzene	12.483	95	24727	36.938	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	73.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050322\
Data File : VY008588.D
Acq On : 03 May 2022 13:45
Operator : KP/MD
Sample : N2605-01
Misc : 5.02g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-04-042822

Quant Time: May 03 16:57:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 01:57:20 2022
Response via : Initial Calibration

