

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050322\
 Data File : VY008579.D
 Acq On : 03 May 2022 09:44
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: May 03 16:54:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 01:57:20 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	116	0.00
2 T	Dichlorodifluoromethane	0.501	0.440	12.2	114	0.00
3 P	Chloromethane	0.582	0.521	10.5	109	0.00
4 C	Vinyl Chloride	0.700	0.622	11.1#	107	0.00
5 T	Bromomethane	0.580	0.481	17.1	106	0.00
6 T	Chloroethane	0.449	0.389	13.4	104	0.00
7 T	Trichlorofluoromethane	0.947	0.858	9.4	109	0.00
8 T	Diethyl Ether	0.303	0.295	2.6	115	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.557	0.520	6.6	114	0.00
10 T	Methyl Iodide	0.690	0.617	10.6	96	0.00
11 T	Tert butyl alcohol	0.058	0.051	12.1	125	0.01
12 CM	1,1-Dichloroethene	0.525	0.502	4.4#	114	0.00
13 T	Acrolein	0.066	0.069	-4.5	126	0.00
14 T	Allyl chloride	0.628	0.624	0.6	116	0.00
15 T	Acrylonitrile	0.138	0.142	-2.9	122	0.00
16 T	Acetone	0.105	0.117	-11.4	129	0.00
17 T	Carbon Disulfide	1.492	1.481	0.7	116	0.00
18 T	Methyl Acetate	0.335	0.301	10.1	120	0.00
19 T	Methyl tert-butyl Ether	1.447	1.382	4.5	113	0.00
20 T	Methylene Chloride	0.799	0.557	30.3#	104	0.00
21 T	trans-1,2-Dichloroethene	0.596	0.574	3.7	114	0.00
22 T	Diisopropyl ether	1.450	1.421	2.0	114	0.00
23 T	Vinyl Acetate	0.947	0.974	-2.9	116	0.00
24 P	1,1-Dichloroethane	0.914	0.878	3.9	113	0.00
25 T	2-Butanone	0.168	0.179	-6.5	125	0.00
26 T	2,2-Dichloropropane	0.876	0.826	5.7	111	0.00
27 T	cis-1,2-Dichloroethene	0.682	0.646	5.3	112	0.00
28 T	Bromochloromethane	0.350	0.314	10.3	106	0.00
29 T	Tetrahydrofuran	0.111	0.117	-5.4	125	0.00
30 C	Chloroform	1.024	0.965	5.8#	110	0.00
31 T	Cyclohexane	0.839	0.798	4.9	116	0.00
32 T	1,1,1-Trichloroethane	0.939	0.873	7.0	109	0.00
33 S	1,2-Dichloroethane-d4	0.486	0.497	-2.3	115	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	116	0.00
35 S	Dibromofluoromethane	0.331	0.348	-5.1	119	0.00
36 T	1,1-Dichloropropene	0.514	0.497	3.3	114	0.00
37 T	Ethyl Acetate	0.264	0.264	0.0	120	0.00
38 T	Carbon Tetrachloride	0.587	0.545	7.2	109	0.00
39 T	Methylcyclohexane	0.679	0.650	4.3	112	0.00
40 TM	Benzene	1.486	1.456	2.0	114	0.00
41 T	Methacrylonitrile	0.116	0.119	-2.6	130	0.00
42 TM	1,2-Dichloroethane	0.430	0.399	7.2	108	0.00
43 T	Isopropyl Acetate	0.480	0.481	-0.2	116	0.00
44 TM	Trichloroethene	0.466	0.435	6.7	111	0.00
45 C	1,2-Dichloropropane	0.347	0.343	1.2#	114	0.00
46 T	Dibromomethane	0.242	0.233	3.7	112	0.00
47 T	Bromodichloromethane	0.527	0.508	3.6	111	0.00
48 T	Methyl methacrylate	0.211	0.209	0.9	116	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050322\
 Data File : VY008579.D
 Acq On : 03 May 2022 09:44
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: May 03 16:54:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 01:57:20 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.003	0.003	0.0	122	0.00
50 S	Toluene-d8	1.172	1.294	-10.4	124	0.00
51 T	4-Methyl-2-Pentanone	0.252	0.260	-3.2	120	0.00
52 CM	Toluene	0.993	0.966	2.7#	114	0.00
53 T	t-1,3-Dichloropropene	0.557	0.547	1.8	113	0.00
54 T	cis-1,3-Dichloropropene	0.619	0.607	1.9	112	0.00
55 T	1,1,2-Trichloroethane	0.337	0.328	2.7	114	0.00
56 T	Ethyl methacrylate	0.426	0.435	-2.1	116	0.00
57 T	1,3-Dichloropropane	0.538	0.533	0.9	115	0.00
58 T	2-Chloroethyl Vinyl ether	0.216	0.238	-10.2	122	0.00
59 T	2-Hexanone	0.172	0.187	-8.7	123	0.00
60 T	Dibromochloromethane	0.422	0.408	3.3	110	0.00
61 T	1,2-Dibromoethane	0.341	0.330	3.2	112	0.00
62 S	4-Bromofluorobenzene	0.430	0.445	-3.5	118	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	112	0.00
64 T	Tetrachloroethene	0.456	0.429	5.9	109	0.00
65 PM	Chlorobenzene	1.204	1.179	2.1	110	0.00
66 T	1,1,1,2-Tetrachloroethane	0.455	0.440	3.3	108	0.00
67 C	Ethyl Benzene	2.026	1.984	2.1#	110	0.00
68 T	m/p-Xylenes	0.828	0.812	1.9	111	0.00
69 T	o-Xylene	0.793	0.781	1.5	111	0.00
70 T	Styrene	1.319	1.314	0.4	111	0.00
71 P	Bromoform	0.302	0.306	-1.3	113	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	108	0.00
73 T	Isopropylbenzene	3.901	3.886	0.4	110	0.00
74 T	N-amyl acetate	0.872	0.952	-9.2	116	0.00
75 P	1,1,2,2-Tetrachloroethane	0.789	0.848	-7.5	119	0.00
76 T	1,2,3-Trichloropropane	0.534	0.530	0.7	117	0.00
77 T	Bromobenzene	1.029	0.996	3.2	109	0.00
78 T	n-propylbenzene	4.539	4.576	-0.8	111	0.00
79 T	2-Chlorotoluene	2.556	2.540	0.6	109	0.00
80 T	1,3,5-Trimethylbenzene	3.223	3.214	0.3	109	0.00
81 T	trans-1,4-Dichloro-2-butene	0.263	0.294	-11.8	119	0.00
82 T	4-Chlorotoluene	2.633	2.651	-0.7	111	0.00
83 T	tert-Butylbenzene	2.886	2.844	1.5	108	0.00
84 T	1,2,4-Trimethylbenzene	3.203	3.177	0.8	109	0.00
85 T	sec-Butylbenzene	4.254	4.209	1.1	109	0.00
86 T	p-Isopropyltoluene	3.612	3.569	1.2	107	0.00
87 T	1,3-Dichlorobenzene	1.990	1.898	4.6	106	0.00
88 T	1,4-Dichlorobenzene	1.963	1.885	4.0	107	0.00
89 T	n-Butylbenzene	3.181	3.193	-0.4	109	0.00
90 T	Hexachloroethane	0.669	0.666	0.4	109	0.00
91 T	1,2-Dichlorobenzene	1.776	1.724	2.9	107	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.120	0.129	-7.5	121	0.00
93 T	1,2,4-Trichlorobenzene	1.133	1.065	6.0	103	0.00
94 T	Hexachlorobutadiene	0.640	0.600	6.3	104	0.00
95 T	Naphthalene	2.273	2.274	-0.0	109	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050322\
Data File : VY008579.D
Acq On : 03 May 2022 09:44
Operator : KP/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: May 03 16:54:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 01:57:20 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	1.006	0.935	7.1	102	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6