

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050324\
 Data File : VY018144.D
 Acq On : 03 May 2024 15:45
 Operator : SY/MD
 Sample : P2362-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 INSIDE

Quant Time: May 06 02:59:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050124S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 16:21:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	238308	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	431584	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	384740	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	154692	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	103796	55.257	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.520%
35) Dibromofluoromethane	7.709	113	125415	50.567	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.140%
50) Toluene-d8	10.178	98	480303	49.605	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.220%
62) 4-Bromofluorobenzene	12.483	95	154361	46.325	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	92.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050324\
Data File : VY018144.D
Acq On : 03 May 2024 15:45
Operator : SY/MD
Sample : P2362-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
INSIDE

Quant Time: May 06 02:59:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050124S.M
Quant Title : SW846 8260
QLast Update : Wed May 01 16:21:20 2024
Response via : Initial Calibration

