

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050324\
 Data File : VY018145.D
 Acq On : 03 May 2024 16:09
 Operator : SY/MD
 Sample : P2362-02
 Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OUTSIDE-FOOTINGS

Quant Time: May 06 02:59:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050124S.M
 Quant Title : SW846 8260
 QLast Update : Wed May 01 16:21:20 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	1834	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	4414	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	3909	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	1023	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	1596	110.402	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	220.800%#
35) Dibromofluoromethane	7.710	113	1587	62.564	ug/l	-0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	125.120%
50) Toluene-d8	10.179	98	4910	49.583	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.160%
62) 4-Bromofluorobenzene	12.483	95	1403	41.169	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	82.340%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050324\
Data File : VY018145.D
Acq On : 03 May 2024 16:09
Operator : SY/MD
Sample : P2362-02
Misc : 5.01g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OUTSIDE-FOOTINGS

Quant Time: May 06 02:59:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050124S.M
Quant Title : SW846 8260
QLast Update : Wed May 01 16:21:20 2024
Response via : Initial Calibration

