

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050422\
 Data File : VY008622.D
 Acq On : 04 May 2022 19:07
 Operator : KP/MD
 Sample : N2619-17
 Misc : 5.39g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SS08-1-2-FT

Quant Time: May 05 04:38:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 01:57:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	60767	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	104657	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	105181	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	48976	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	30543	51.714	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	103.420%
35) Dibromofluoromethane	7.716	113	33589	48.503	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.000%
50) Toluene-d8	10.179	98	123279	50.247	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.500%
62) 4-Bromofluorobenzene	12.483	95	41895	46.589	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	93.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050422\
Data File : VY008622.D
Acq On : 04 May 2022 19:07
Operator : KP/MD
Sample : N2619-17
Misc : 5.39g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SS08-1-2-FT

Quant Time: May 05 04:38:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 01:57:20 2022
Response via : Initial Calibration

