

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050422\
 Data File : VY008633.D
 Acq On : 05 May 2022 00:30
 Operator : KP/MD
 Sample : N2619-19
 Misc : 5.51g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 38 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SS09-1-2-FT

Quant Time: May 05 04:42:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 01:57:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	131136	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	227290	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	233403	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	110564	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	63835	50.085	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 100.160%	
35) Dibromofluoromethane	7.715	113	72227	48.024	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 96.040%	
50) Toluene-d8	10.178	98	270898	50.841	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 101.680%	
62) 4-Bromofluorobenzene	12.483	95	96401	49.362	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 98.720%	
Target Compounds						
16) Acetone	3.948	43	3569	12.996	ug/l	Qvalue # 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050422\
Data File : VY008633.D
Acq On : 05 May 2022 00:30
Operator : KP/MD
Sample : N2619-19
Misc : 5.51g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 38 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SS09-1-2-FT

Quant Time: May 05 04:42:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 01:57:20 2022
Response via : Initial Calibration

