

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050422\
 Data File : VY008640.D
 Acq On : 05 May 2022 03:14
 Operator : KP/MD
 Sample : N2665-03
 Misc : 5.16g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 45 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-8,6

Quant Time: May 05 04:43:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 01:57:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	117503	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	200401	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	201866	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	99655	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	56131	49.150	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.300%
35) Dibromofluoromethane	7.716	113	61839	46.634	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	93.260%
50) Toluene-d8	10.179	98	240275	51.145	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.280%
62) 4-Bromofluorobenzene	12.483	95	87304	50.702	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050422\
Data File : VY008640.D
Acq On : 05 May 2022 03:14
Operator : KP/MD
Sample : N2665-03
Misc : 5.16g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 45 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-8,6

Quant Time: May 05 04:43:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 01:57:20 2022
Response via : Initial Calibration

