

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050422\
 Data File : VY008643.D
 Acq On : 05 May 2022 04:24
 Operator : KP/MD
 Sample : N2665-06
 Misc : 4.13g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 48 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-4,11

Quant Time: May 05 06:52:57 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 01:57:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	111743	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	180986	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	183187	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	92435	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	53544	49.301	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.600%
35) Dibromofluoromethane	7.716	113	56574	47.240	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	94.480%
50) Toluene-d8	10.179	98	212746	50.143	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	100.280%
62) 4-Bromofluorobenzene	12.483	95	79859	51.353	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	102.700%
Target Compounds						
					Qvalue	
68) m/p-Xylenes	11.697	106	5165	1.703	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050422\
Data File : VY008643.D
Acq On : 05 May 2022 04:24
Operator : KP/MD
Sample : N2665-06
Misc : 4.13g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 48 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-4,11

Quant Time: May 05 06:52:57 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 01:57:20 2022
Response via : Initial Calibration

