

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050422\
 Data File : VY008647.D
 Acq On : 05 May 2022 05:58
 Operator : KP/MD
 Sample : N2665-10
 Misc : 4.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-10,8

Quant Time: May 05 06:53:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 01:57:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	52969	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	103567	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	109913	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	96015	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	35137	68.251	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	136.500%
35) Dibromofluoromethane	7.715	113	34636	50.541	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.080%
50) Toluene-d8	10.178	98	129978	53.535	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.080%
62) 4-Bromofluorobenzene	12.483	95	51948	58.376	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	116.760%
Target Compounds						Qvalue
30) Chloroform	7.496	83	1241	1.144	ug/l	98
31) Cyclohexane	7.776	56	1312	1.476	ug/l #	25

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050422\
Data File : VY008647.D
Acq On : 05 May 2022 05:58
Operator : KP/MD
Sample : N2665-10
Misc : 4.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-10,8

Quant Time: May 05 06:53:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 01:57:20 2022
Response via : Initial Calibration

