

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050422\
 Data File : VY008648.D
 Acq On : 05 May 2022 06:21
 Operator : KP/MD
 Sample : N2665-11
 Misc : 5.77g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 53 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-10,14

Quant Time: May 05 06:54:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 01:57:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.783	168	51260	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	100232	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	103452	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	47180	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	34755	69.760	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 139.520%	
35) Dibromofluoromethane	7.716	113	32460	48.942	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 97.880%	
50) Toluene-d8	10.179	98	127220	54.143	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 108.280%	
62) 4-Bromofluorobenzene	12.483	95	48210	55.978	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 111.960%	
Target Compounds						
					Qvalue	
30) Chloroform	7.508	83	1302	1.240	ug/l #	71
31) Cyclohexane	7.789	56	1231	1.432	ug/l #	46

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050422\
Data File : VY008648.D
Acq On : 05 May 2022 06:21
Operator : KP/MD
Sample : N2665-11
Misc : 5.77g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 53 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-10,14

Quant Time: May 05 06:54:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 01:57:20 2022
Response via : Initial Calibration

