

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050422\
 Data File : VY008652.D
 Acq On : 05 May 2022 07:55
 Operator : KP/MD
 Sample : N2650-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ETGI-349

Quant Time: May 05 09:06:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 01:57:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	0.000	168	0	0.000	ug/l	-7.79
34) 1,4-Difluorobenzene	0.000	114	0	0.000	ug/l	-8.69
63) Chlorobenzene-d5	11.642	117	6521	50.000	ug/l	0.15
72) 1,4-Dichlorobenzene-d4	13.416	152	773	50.000	ug/l	-0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.100	65	19	0.000	ug/l	-0.04
Spiked Amount	50.000	Range	50 - 163	Recovery	=	0.000%#
35) Dibromofluoromethane	0.000	113	0	0.000	ug/l	
Spiked Amount	50.000	Range	54 - 147	Recovery	=	0.000%#
50) Toluene-d8	10.331	98	77	0.000	ug/l	0.15
Spiked Amount	50.000	Range	49 - 140	Recovery	=	0.000%#
62) 4-Bromofluorobenzene	12.556	95	4245	0.000	ug/l	0.07
Spiked Amount	50.000	Range	25 - 144	Recovery	=	0.000%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050422\
Data File : VY008652.D
Acq On : 05 May 2022 07:55
Operator : KP/MD
Sample : N2650-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 57 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ETGI-349

Quant Time: May 05 09:06:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 01:57:20 2022
Response via : Initial Calibration

