

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050422\
 Data File : VY008609.D
 Acq On : 04 May 2022 14:03
 Operator : KP/MD
 Sample : N2619-02
 Misc : 5.82g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB01-13-14-FT

Quant Time: May 05 04:35:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
 Quant Title : SW846 8260
 QLast Update : Tue May 03 01:57:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	132105	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	223915	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	231958	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	116738	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	65904	51.329	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	102.660%
35) Dibromofluoromethane	7.716	113	71350	48.156	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	96.320%
50) Toluene-d8	10.179	98	268191	51.092	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.180%
62) 4-Bromofluorobenzene	12.483	95	97715	50.789	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.580%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050422\
Data File : VY008609.D
Acq On : 04 May 2022 14:03
Operator : KP/MD
Sample : N2619-02
Misc : 5.82g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB01-13-14-FT

Quant Time: May 05 04:35:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
Quant Title : SW846 8260
QLast Update : Tue May 03 01:57:20 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY008609.D
Acq: 04 May 2022 14:03




Instrument :
MSVOA_Y
ClientSampleId :
SB01-13-14-FT


Tgt Ion:168 Resp: 132105
Ion Ratio Lower Upper
168 100
99 46.6 46.9 70.3#



#33
1,2-Dichloroethane-d4
Concen: 51.329 ug/l
RT: 8.143 min Scan# 1037
Delta R.T. 0.000 min
Lab File: VY008609.D
Acq: 04 May 2022 14:03

Tgt Ion: 65 Resp: 65904
Ion Ratio Lower Upper
65 100
67 55.8 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. 0.000 min

Lab File: VY008609.D

Acq: 04 May 2022 14:03

Instrument :

MSVOA_Y

ClientSampleId :

SB01-13-14-FT





Tgt Ion:114 Resp: 223915

Ion Ratio Lower Upper

114 100

63 16.1 0.0 43.2

88 13.4 0.0 31.8



#35

Dibromofluoromethane

Concen: 48.156 ug/l

RT: 7.716 min Scan# 967

Delta R.T. 0.000 min

Lab File: VY008609.D

Acq: 04 May 2022 14:03




Tgt Ion:113 Resp: 71350


Ion Ratio Lower Upper

113 100

111 102.1 82.6 124.0

192 21.9 17.6 26.4









#50
Toluene-d8
Concen: 51.092 ug/l
RT: 10.179 min Scan# 1371
Delta R.T. 0.000 min
Lab File: VY008609.D
Acq: 04 May 2022 14:03




Instrument :
MSVOA_Y
ClientSampleId :
SB01-13-14-FT


Tgt Ion: 98 Resp: 268191
Ion Ratio Lower Upper
98 100
100 64.8 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 50.789 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY008609.D
Acq: 04 May 2022 14:03

Tgt Ion: 95 Resp: 97715
Ion Ratio Lower Upper
95 100
174 98.3 0.0 177.0
176 94.0 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY008609.D
Acq: 04 May 2022 14:03

Instrument :
MSVOA_Y
ClientSampleId :
SB01-13-14-FT

Tgt Ion:117 Resp: 231958
Ion Ratio Lower Upper
117 100
82 50.5 43.0 64.4
119 31.5 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY008609.D
Acq: 04 May 2022 14:03

Tgt Ion:152 Resp: 116738
Ion Ratio Lower Upper
152 100
115 54.4 28.2 84.7
150 157.0 0.0 347.6

