



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050422\
 Data File : VY008600.D
 Acq On : 04 May 2022 09:01
 Operator : KP/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050222S.M
 Title : SW846 8260
 Last Update : Tue May 03 01:57:20 2022



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1741

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	16.5	7169	PASS
75	95	30	60	49.5	21464	PASS
95	95	100	100	100.0	43395	PASS
96	95	5	9	6.8	2969	PASS
173	174	0.00	2	0.7	313	PASS
174	95	50	100	97.3	42235	PASS
175	174	5	9	7.4	3144	PASS
176	174	95	101	97.7	41253	PASS
177	176	5	9	7.2	2982	PASS