

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050525\
 Data File : VY022164.D
 Acq On : 05 May 2025 17:05
 Operator : SY/MD
 Sample : Q1956-06
 Misc : 6.15g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB91-3-4

Quant Time: May 06 05:23:01 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	214691	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	422744	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	385176	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	143859	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	112650	56.806	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 113.620%	
35) Dibromofluoromethane	7.634	113	143007	51.204	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.400%	
50) Toluene-d8	10.109	98	519895	49.377	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 98.760%	
62) 4-Bromofluorobenzene	12.408	95	150082	42.342	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 84.680%	
Target Compounds						Qvalue
20) Methylene Chloride	4.616	84	10335	4.122	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050525\
Data File : VY022164.D
Acq On : 05 May 2025 17:05
Operator : SY/MD
Sample : Q1956-06
Misc : 6.15g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB91-3-4

Quant Time: May 06 05:23:01 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

