

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050625\
 Data File : VY022174.D
 Acq On : 06 May 2025 11:29
 Operator : SY/MD
 Sample : Q1961-01
 Misc : 5.29g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 EO-01-050525

Quant Time: May 07 04:33:20 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	94987	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.622	114	153250	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	123852	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	36167	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	43218	49.258	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	98.520%
35) Dibromofluoromethane	7.634	113	55798	55.112	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.220%
50) Toluene-d8	10.109	98	178874	46.863	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	93.720%
62) 4-Bromofluorobenzene	12.408	95	40727	31.696	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	63.400%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050625\
Data File : VY022174.D
Acq On : 06 May 2025 11:29
Operator : SY/MD
Sample : Q1961-01
Misc : 5.29g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
EO-01-050525

Quant Time: May 07 04:33:20 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

