

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050625\
 Data File : VY022177.D
 Acq On : 06 May 2025 12:40
 Operator : SY/MD
 Sample : Q1960-02
 Misc : 4.48g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 72-11933

Quant Time: May 07 04:34:27 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	211599	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.615	114	419403	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	380157	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	145755	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.060	65	118561	60.660	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.320%
35) Dibromofluoromethane	7.634	113	143997	51.969	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.940%
50) Toluene-d8	10.109	98	511849	49.000	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.000%
62) 4-Bromofluorobenzene	12.407	95	143933	40.931	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	81.860%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050625\
Data File : VY022177.D
Acq On : 06 May 2025 12:40
Operator : SY/MD
Sample : Q1960-02
Misc : 4.48g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
72-11933

Quant Time: May 07 04:34:27 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

