

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050625\
 Data File : VY022170.D
 Acq On : 06 May 2025 09:14
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: May 07 04:30:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	75	0.00
2 T	Dichlorodifluoromethane	0.426	0.377	11.5	72	0.00
3 P	Chloromethane	0.607	0.572	5.8	71	0.01
4 C	Vinyl Chloride	0.745	0.766	-2.8#	78	0.01
5 T	Bromomethane	0.642	0.588	8.4	74	0.00
6 T	Chloroethane	0.508	0.522	-2.8	79	0.00
7 T	Trichlorofluoromethane	0.983	0.948	3.6	74	0.01
8 T	Diethyl Ether	0.256	0.240	6.3	72	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.533	0.526	1.3	80	0.01
10 T	Methyl Iodide	0.617	0.593	3.9	70	0.00
11 T	Tert butyl alcohol	0.025	0.025	0.0	71	-0.01
12 CM	1,1-Dichloroethene	0.497	0.479	3.6#	76	0.00
13 T	Acrolein	0.046	0.017	63.0#	30#	0.00
14 T	Allyl chloride	0.579	0.597	-3.1	80	0.00
15 T	Acrylonitrile	0.094	0.095	-1.1	76	0.00
16 T	Acetone	0.070	0.077	-10.0	91	0.00
17 T	Carbon Disulfide	1.585	1.477	6.8	74	0.01
18 T	Methyl Acetate	0.216	0.248	-14.8	83	0.00
19 T	Methyl tert-butyl Ether	1.113	1.119	-0.5	74	0.00
20 T	Methylene Chloride	0.584	0.527	9.8	74	0.00
21 T	trans-1,2-Dichloroethene	0.557	0.556	0.2	78	0.01
22 T	Diisopropyl ether	1.288	1.394	-8.2	81	0.00
23 T	Vinyl Acetate	0.694	0.744	-7.2	78	0.00
24 P	1,1-Dichloroethane	0.893	0.924	-3.5	81	0.00
25 T	2-Butanone	0.105	0.113	-7.6	80	0.00
26 T	2,2-Dichloropropane	0.782	0.830	-6.1	83	0.00
27 T	cis-1,2-Dichloroethene	0.622	0.630	-1.3	77	0.00
28 T	Bromochloromethane	0.347	0.350	-0.9	76	0.00
29 T	Tetrahydrofuran	0.068	0.074	-8.8	76	0.00
30 C	Chloroform	0.990	1.019	-2.9#	79	0.00
31 T	Cyclohexane	0.763	0.776	-1.7	82	0.00
32 T	1,1,1-Trichloroethane	0.896	0.908	-1.3	79	0.00
33 S	1,2-Dichloroethane-d4	0.462	0.462	0.0	74	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	72	0.00
35 S	Dibromofluoromethane	0.330	0.334	-1.2	73	0.00
36 T	1,1-Dichloropropene	0.444	0.476	-7.2	81	0.00
37 T	Ethyl Acetate	0.159	0.174	-9.4	78	0.00
38 T	Carbon Tetrachloride	0.530	0.565	-6.6	80	0.00
39 T	Methylcyclohexane	0.558	0.603	-8.1	79	0.00
40 TM	Benzene	1.387	1.468	-5.8	79	0.00
41 T	Methacrylonitrile	0.087	0.091	-4.6	72	0.00
42 TM	1,2-Dichloroethane	0.343	0.366	-6.7	79	0.00
43 T	Isopropyl Acetate	0.306	0.328	-7.2	76	0.00
44 TM	Trichloroethene	0.384	0.400	-4.2	79	0.00
45 C	1,2-Dichloropropane	0.307	0.336	-9.4#	81	0.00
46 T	Dibromomethane	0.192	0.203	-5.7	78	0.00
47 T	Bromodichloromethane	0.480	0.513	-6.9	79	0.00
48 T	Methyl methacrylate	0.141	0.158	-12.1	76	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050625\
 Data File : VY022170.D
 Acq On : 06 May 2025 09:14
 Operator : SY/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: May 07 04:30:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	76	0.00
50 S	Toluene-d8	1.245	1.283	-3.1	73	0.00
51 T	4-Methyl-2-Pentanone	0.160	0.183	-14.4	77	0.00
52 CM	Toluene	0.898	0.976	-8.7#	79	0.00
53 T	t-1,3-Dichloropropene	0.411	0.453	-10.2	78	0.00
54 T	cis-1,3-Dichloropropene	0.489	0.525	-7.4	78	0.00
55 T	1,1,2-Trichloroethane	0.256	0.276	-7.8	78	0.00
56 T	Ethyl methacrylate	0.293	0.320	-9.2	73	0.00
57 T	1,3-Dichloropropane	0.411	0.439	-6.8	78	0.00
58 T	2-Chloroethyl Vinyl ether	0.143	0.157	-9.8	71	0.00
59 T	2-Hexanone	0.106	0.123	-16.0	77	0.00
60 T	Dibromochloromethane	0.355	0.376	-5.9	77	0.00
61 T	1,2-Dibromoethane	0.242	0.256	-5.8	77	0.00
62 S	4-Bromofluorobenzene	0.419	0.436	-4.1	74	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	72	0.00
64 T	Tetrachloroethene	0.472	0.490	-3.8	78	0.00
65 PM	Chlorobenzene	1.118	1.178	-5.4	79	0.00
66 T	1,1,1,2-Tetrachloroethane	0.394	0.414	-5.1	77	0.00
67 C	Ethyl Benzene	1.829	2.031	-11.0#	80	0.00
68 T	m/p-Xylenes	0.728	0.817	-12.2	80	0.00
69 T	o-Xylene	0.671	0.744	-10.9	78	0.00
70 T	Styrene	1.126	1.278	-13.5	79	0.00
71 P	Bromoform	0.229	0.239	-4.4	75	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	75	0.00
73 T	Isopropylbenzene	3.341	3.536	-5.8	80	0.00
74 T	N-amyl acetate	0.574	0.593	-3.3	75	0.00
75 P	1,1,2,2-Tetrachloroethane	0.563	0.578	-2.7	79	0.00
76 T	1,2,3-Trichloropropane	0.428	0.395	7.7	73	0.00
77 T	Bromobenzene	0.844	0.839	0.6	76	0.00
78 T	n-propylbenzene	3.978	4.340	-9.1	82	0.00
79 T	2-Chlorotoluene	2.314	2.445	-5.7	81	0.00
80 T	1,3,5-Trimethylbenzene	2.752	2.992	-8.7	80	0.00
81 T	trans-1,4-Dichloro-2-butene	0.178	0.184	-3.4	78	0.00
82 T	4-Chlorotoluene	2.405	2.563	-6.6	81	0.00
83 T	tert-Butylbenzene	2.468	2.618	-6.1	79	0.00
84 T	1,2,4-Trimethylbenzene	2.753	3.014	-9.5	81	0.00
85 T	sec-Butylbenzene	3.612	3.931	-8.8	82	0.00
86 T	p-Isopropyltoluene	3.067	3.352	-9.3	82	0.00
87 T	1,3-Dichlorobenzene	1.714	1.771	-3.3	81	0.00
88 T	1,4-Dichlorobenzene	1.698	1.748	-2.9	81	0.00
89 T	n-Butylbenzene	2.742	3.028	-10.4	83	0.00
90 T	Hexachloroethane	0.674	0.700	-3.9	83	0.00
91 T	1,2-Dichlorobenzene	1.495	1.539	-2.9	80	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.090	0.090	0.0	81	0.00
93 T	1,2,4-Trichlorobenzene	0.847	0.851	-0.5	78	0.00
94 T	Hexachlorobutadiene	0.521	0.519	0.4	81	0.00
95 T	Naphthalene	1.408	1.405	0.2	74	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050625\
Data File : VY022170.D
Acq On : 06 May 2025 09:14
Operator : SY/MD
Sample : VSTDCCC050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_Y
LabSampleId :
VSTDCCC050

Quant Time: May 07 04:30:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.731	0.731	0.0	78	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6