

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050725\
 Data File : VY022187.D
 Acq On : 07 May 2025 12:27
 Operator : SY/MD
 Sample : Q1962-01
 Misc : 5.53g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HD-01-050525

Quant Time: May 08 01:41:48 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 23 02:30:30 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.707	168	153931	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.616	114	287873	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	252442	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	89330	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	88680	62.370	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	124.740%
35) Dibromofluoromethane	7.634	113	105919	55.693	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	111.380%
50) Toluene-d8	10.109	98	351064	48.963	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	97.920%
62) 4-Bromofluorobenzene	12.408	95	89957	37.269	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	74.540%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY050725\
Data File : VY022187.D
Acq On : 07 May 2025 12:27
Operator : SY/MD
Sample : Q1962-01
Misc : 5.53g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HD-01-050525

Quant Time: May 08 01:41:48 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y042225S.M
Quant Title : SW846 8260
QLast Update : Wed Apr 23 02:30:30 2025
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.707 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY022187.D
Acq: 07 May 2025 12:27




Instrument :
MSVOA_Y
ClientSampleId :
HD-01-050525

Tgt Ion:168 Resp: 153931
Ion Ratio Lower Upper
168 100
99 56.5 43.1 64.7



#33
1,2-Dichloroethane-d4
Concen: 62.370 ug/l
RT: 8.061 min Scan# 1040
Delta R.T. -0.000 min
Lab File: VY022187.D
Acq: 07 May 2025 12:27

Tgt Ion: 65 Resp: 88680
Ion Ratio Lower Upper
65 100
67 52.2 0.0 105.8









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.616 min Scan# 1131
Delta R.T. -0.000 min
Lab File: VY022187.D
Acq: 07 May 2025 12:27




Instrument :
MSVOA_Y
ClientSampleId :
HD-01-050525

Tgt Ion:114 Resp: 287873
Ion Ratio Lower Upper
114 100
63 17.9 0.0 35.4
88 14.4 0.0 28.2



#35
Dibromofluoromethane
Concen: 55.693 ug/l
RT: 7.634 min Scan# 970
Delta R.T. -0.000 min
Lab File: VY022187.D
Acq: 07 May 2025 12:27

Tgt Ion:113 Resp: 105919
Ion Ratio Lower Upper
113 100
111 100.8 81.8 122.8
192 19.7 15.6 23.4








#50
Toluene-d8
Concen: 48.963 ug/l
RT: 10.109 min Scan# 1376
Delta R.T. 0.006 min
Lab File: VY022187.D
Acq: 07 May 2025 12:27




Instrument :
MSVOA_Y
ClientSampleId :
HD-01-050525


Tgt Ion: 98 Resp: 351064
Ion Ratio Lower Upper
98 100
100 63.5 51.6 77.4



#62
4-Bromofluorobenzene
Concen: 37.269 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. 0.006 min
Lab File: VY022187.D
Acq: 07 May 2025 12:27

Tgt Ion: 95 Resp: 89957
Ion Ratio Lower Upper
95 100
174 86.8 0.0 179.0
176 84.7 0.0 171.8





#63

Chlorobenzene-d5

Concen: 50.000 ug/l

RT: 11.414 min Scan# 117

Delta R.T. -0.000 min

Lab File: VY022187.D

Acq: 07 May 2025 12:27

Instrument :

MSVOA_Y

ClientSampleId :

HD-01-050525





Tgt Ion:117 Resp: 252442

Ion Ratio Lower Upper

117 100

82 54.1 42.1 63.1

119 33.0 25.7 38.5



#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.346 min Scan# 1907

Delta R.T. -0.000 min

Lab File: VY022187.D

Acq: 07 May 2025 12:27

Tgt Ion:152 Resp: 89330

Ion Ratio Lower Upper

152 100

115 56.0 28.0 84.0

150 156.8 0.0 345.6

