

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051022\
 Data File : VY008678.D
 Acq On : 10 May 2022 14:07
 Operator : KP/MD
 Sample : N2760-01
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-34

Manual Integrations
 APPROVED

Reviewed By : John Carlone 05/11/2022
 Supervised By : Mahesh Dadoda 05/11/2022

Quant Time: May 11 04:30:44 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M
 Quant Title : SW846 8260
 QLast Update : Mon May 09 19:21:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	2049	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	3434	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	3262	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	954	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.130	65	1491m	84.924	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 169.840%#	
35) Dibromofluoromethane	7.710	113	1231	58.171	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 116.340%	
50) Toluene-d8	10.179	98	3970	53.236	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 106.480%	
62) 4-Bromofluorobenzene	12.483	95	1196	44.578	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 89.160%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051022\
Data File : VY008678.D
Acq On : 10 May 2022 14:07
Operator : KP/MD
Sample : N2760-01
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-34

Manual Integrations
APPROVED

Reviewed By : John Carlone 05/11/2022
Supervised By : Mahesh Dadoda 05/11/2022

Quant Time: May 11 04:30:44 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M
Quant Title : SW846 8260
QLast Update : Mon May 09 19:21:20 2022
Response via : Initial Calibration

