

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051022\
 Data File : VY008686.D
 Acq On : 10 May 2022 17:44
 Operator : KP/MD
 Sample : N2762-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB02

Quant Time: May 11 04:33:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M
 Quant Title : SW846 8260
 QLast Update : Mon May 09 19:21:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) Pentafluorobenzene	7.795	168	2515	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	4311	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	3496	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	849	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	1361	63.156	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.320%
35) Dibromofluoromethane	7.710	113	1493	56.199	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	112.400%
50) Toluene-d8	10.179	98	4587	48.997	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	98.000%
62) 4-Bromofluorobenzene	12.483	95	1126	33.431	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	66.860%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051022\
Data File : VY008686.D
Acq On : 10 May 2022 17:44
Operator : KP/MD
Sample : N2762-01
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB02

Quant Time: May 11 04:33:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M
Quant Title : SW846 8260
QLast Update : Mon May 09 19:21:20 2022
Response via : Initial Calibration

