

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051022\
 Data File : VY008690.D
 Acq On : 10 May 2022 19:17
 Operator : KP/MD
 Sample : N2762-05
 Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB15

Quant Time: May 11 04:34:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M
 Quant Title : SW846 8260
 QLast Update : Mon May 09 19:21:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	40910	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	62179	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	51879	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	20935	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	16822	47.989	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	95.980%
35) Dibromofluoromethane	7.716	113	19188	50.076	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.160%
50) Toluene-d8	10.179	98	69754	51.659	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	103.320%
62) 4-Bromofluorobenzene	12.483	95	18846	38.794	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	77.580%
Target Compounds						
16) Acetone	3.936	43	1644	19.879	ug/l	# 85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051022\
Data File : VY008690.D
Acq On : 10 May 2022 19:17
Operator : KP/MD
Sample : N2762-05
Misc : 5.03g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB15

Quant Time: May 11 04:34:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M
Quant Title : SW846 8260
QLast Update : Mon May 09 19:21:20 2022
Response via : Initial Calibration

