

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051022\
 Data File : VY008693.D
 Acq On : 10 May 2022 20:28
 Operator : KP/MD
 Sample : N2762-08
 Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB18

Quant Time: May 11 04:35:20 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M
 Quant Title : SW846 8260
 QLast Update : Mon May 09 19:21:20 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	219256	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.685	114	363035	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	331033	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	153057	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	102947	54.797	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	= 109.600%		
35) Dibromofluoromethane	7.716	113	113897	50.911	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	= 101.820%		
50) Toluene-d8	10.179	98	423085	53.666	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	= 107.340%		
62) 4-Bromofluorobenzene	12.483	95	133741	47.153	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	= 94.300%		
Target Compounds						
						Qvalue
16) Acetone	3.936	43	6698	15.112	ug/l	# 64
20) Methylene Chloride	4.710	84	16536	6.143	ug/l	# 70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY051022\
Data File : VY008693.D
Acq On : 10 May 2022 20:28
Operator : KP/MD
Sample : N2762-08
Misc : 5.04g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB18

Quant Time: May 11 04:35:20 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y050922S.M
Quant Title : SW846 8260
QLast Update : Mon May 09 19:21:20 2022
Response via : Initial Calibration

